



**CALIBRATION REPORT**

Report No : HBMS/CAL/22-23		Calibration Date : 21.03.2023	
		Calibration Valid: 20.09.2023	
<b>I.1.CUSTOMER DETAILS:</b>			
Name and Address of the organization		<b>M/s, UPGRADED PRIMARY HEALTH CENTRE, THIRUMARUGAL, NAGAPATTINAM.</b>	
<b>I.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>			
EQUIPMENT NAME		MICRO PIPETTE	
i	Manufactured by	THERMO SCIENTIFIC	
ii	Model	FINNEPIPEPTE F3	
iii	Value	10 µl -100 µl	
iv	Serial No	RW 09907	


**Calibration standards used :**

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

**Calibration test Report.**

Number of measurement	Channel I (Volume)		
	10µl	50µl	100µl
1	10.03	50.06	100.14
2	10.04	50.09	100.12
3	10.05	50.02	100.10
4	10.09	50.08	100.08
5	10.00	49.94	100.09
6	10.08	50.07	100.12

For Heart Beat Medical Systems,  
(Authorized signatory)



No. 23, 2nd Floor, jayasimmapuram,  
Pappanaikenpalayam, Coimbatore – 37.

99400 28355, 97512 06106  
88833 15662

hbmedicalsystem@gmail.com  
hbmssales@gmail.com





CALIBRATION REPORT	
Report No : HBMS/CAL/22-23	Calibration Date : 21.03.2023
	Calibration Valid: 20.09.2023
<b>1.1.CUSTOMER DETAILS:</b>	
Name and Address of the organization	M/s, UPGRADED PRIMARY HEALTH CENTRE, THIRUMARUGAL, NAGAPATTINAM.
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>	
EQUIPMENT NAME	
i   Manufactured by	MICRO PIPETTE THEPETTENANN
ii   Model	PFACT
iii   Value	10 µl -100 µl
iv   Serial No	242415


**Calibration standards used :**

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

**Calibration test Report.**

Number of measurement	Channel 1 (Volume)		
	10µl	50µl	100µl
1	10.02	50.04	100.06
2	10.03	50.08	100.04
3	10.05	50.01	100.08
4	10.08	50.07	100.07
5	10.01	49.93	100.05
6	10.09	50.05	100.09

For Heart Beat Medical Systems,  
 (Authorized signatory)



No. 23, 2nd Floor, jayasimmapuram,  
 Pappanaikenpalayam, Coimbatore - 37.

99400 28355, 97512 06106  
 88833 15662

hbmedicalsystem@gmail.com  
 hbmsales@gmail.com

### CALIBRATION REPORT

Report No: HBMS/CAL/22-23		Calibration Date: 21.03.2023
		Calibration Valid: 20.09.2023
<b>1.1. CUSTOMER DETAILS:</b>		
Name and Address of the organization		<b>M/s. UPGRADED PRIMARY HEALTH CENTRE, THIRUMARUGAL, NAGAPATTINAM.</b>
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>		
EQUIPMENT NAME		PIPETTE
I	Manufactured by	THERMO SCIENTIFIC
II	Model	FINNEPIPETTE F3
III	Value	100µl - 1000µl
IV	Serial No	RW 12645

**Details of the machine:** machine calibrated and checked output.

Number of measurement	Channel 1 (Volume)		
	100µl	500µl	1000µl
1	100.05	500.08	999.97
2	100.06	500.07	1000.02
3	100.01	500.05	999.96
4	99.97	499.99	1000.02
5	100.01	500.08	1000.09
6	100.02	499.95	1000.04

**For Heart Beat Medical Systems,**

*(Signature)*  
**(Authorized signatory)**

No. 23, 2nd Floor, jayasimmapuram,  
Pappanaikenpalayam, Coimbatore – 37.

99400 28355, 97512 06106

88833 15662

hbmedicalsystem@gmail.com

hbmssales@gmail.com