

ST-200aQua CALIBRATION CERTIFICATE

CUSTOMERS NAME: **INTEGRO HOSPITAL**
.....

ADDRESS: **D. No:12-2-725/2, P&T Co. Op Housing Society, Rethibowli, Hyderabad.**
.....

CONTACT NO:
.....

SERIAL NO OF THE INSTRUMENT: **STAQW-3499**
.....

CALIBRATION DATE: **03/01/2023**
.....

VALID TILL: **02/01/2024**
.....

CALIBRATION DONE BY: **Mr. KIRAN DATE**
.....

Calibration Slopes:				
S.No.	Na ⁺	K ⁺	iCa	Cl ⁻
1	59.14	55.17	30.99	43.75
2	58.19	55.69	31.01	44.01
3	59.20	55.78	30.91	43.79
Mean	58.84	55.54	30.97	43.82
Ranges	30-100mV/dec	30-100 mV/dec	15 - 50 mV/dec	30 to 65 mV/dec
Status	O.K	O.K	O.K	O.K

Sensa Core Medical Instrumentation Pvt. Ltd.

Plot, No: 3, Export Promotion Industrial Park, Pashamylaram, Medak (Dist), 502307, T.S. INDIA.
Tel: +08455 - 223400, www.sensacore.com

SENSA CORE
Healthcare and Diagnostic Products

Quality Control 1:		Lot:- UQC2210A-L1			Exp: Mar 2024
S.No.	Na+	K+	iCa	Cl ⁻	
1	120.3	1.92	1.76	79.78	
2	121.5	1.98	1.80	81.01	
3	121.3	1.96	1.83	81.70	
Mean	121.0	1.95	1.79	80.83	
Ranges	113-128 mmol/L	1.5-2.5mmol/L	1.25-2.25mmol/L	76-90mmol/L	
Status	O.K	O.K	O.K	O.K	
Quality Control 2:		Lot:- UQC2210A-L2			Exp: Mar 2024
S.No.	Na+	K+	iCa	Cl ⁻	
1	141.07	3.83	1.20	102.5	
2	142.19	3.89	1.25	101.3	
3	141.13	3.95	1.23	101.9	
Mean	141.46	3.89	1.22	101.9	
Ranges	132-153mmol/L	3.5-4.7mmol/L	0.62-1.62mmol/L	95-111mmol/L	
Status	O.K	O.K	O.K	O.K	
Quality Control 3:		Lot:- UQC2210A-L3			Exp: Mar 2024
S.No.	Na+	K+	iCa	Cl ⁻	
1	158.91	5.89	0.79	120.4	
2	160.09	5.93	0.83	121.2	
3	159.09	5.90	0.85	120.9	
Mean	159.36	5.90	0.82	120.8	
Ranges	152-164mmol/L	5.4-6.8mmol/L	0.75-1.15mmol/L	116-128mmol/L	
Status	O.K	O.K	O.K	O.K	



Signature of Engineer

Signature of Doctor *[Signature]* 03/01/23

DATE: 03/01/2023

DATE:

Sensa Core Medical Instrumentation Pvt. Ltd.

Plot, No: 3, Export Promotion Industrial Park, Pashamylaram, Medak (Dist), 502307, T.S. INDIA.

Tel: +08455 - 223400, www.sensacore.com