

# CALIBRATION REPORT

Refrigerators, Laboratory(Lab refrigerator)



Report No:	CAL-AR-2023-01-1078	Date: 09-01-2023	Page 01 of 02
------------	---------------------	------------------	---------------

Customer Details			
Health Facility	Gyati Takka GH, Ziro		
State	Arunachal Pradesh	District	Lower Subansiri

Medical Equipment Details			
ME Number	ARZ1LOSDH0070379	Manufacturer	Hair
ME Name	Refrigerators, Laboratory(Lab refrigerator)	Model	NA
Department	ICTC, Near OPD	Serial No	NA
Calibrated at	On Site	Condition of ME	Good
Date of Calibration	02-01-2023	Due for Calibration	01-01-2024

Environmental Conditions			
Temperature in °C	20°C	Humidity in % Rh	58%

Details of Master Equipment Used for Calibration						
No	Equipment ID	Description	Manufacturer	Model	Serial No.	Calibration Validity
1	TRI-TE-0381	DIGITAL THERMOMETER	METRAVI	DTM-901	210654963	07-05-2023

Performance Results						
No	Parameter	Set Temp	Unit	Measured Value	Acceptable Range	Remarks
1	Temperature	2°C	°C	2.1	1.5 to 2.5°C	Pass
2	Temperature	4°C	°C	4.1	3.5 to 4.5°C	Pass
3	Temperature	6°C	°C	6.2	5.5 to 6.5°C	Pass

# CALIBRATION REPORT

Refrigerators, Laboratory(Lab refrigerator)



## Conditions of Issue of report

1. The Standards used are traceable to National standards.
2. This report should not be reproduced except in full without written approval from Trimed Solutions India Private Limited.

**Calibration Status**

Passed

Failed

**Calibrated By**

*Pathik Maity*

Pathik Maity

**Authorised Signatory**

*Rahul Kumar Kashyap*

Rahul Kumar Kashyap