

KIRLOSKAR TECHNOLOGIES (P) LTD.
(Quality Assessment Services)



B-58, 1st Floor, Defence Colony, Bhisham Pitamah Marg, New Delhi-110024, INDIA
Phone: 011-24331368,24331369,41403967; E-mail: gianenender.verma@kirloskarmedical.com

Calibration Certificate

Page 1 of 2

Certificate Number:	221028.8.2	ULR No.	N/A			
Calibration Date:	28-10-2022	Next Recommended Cal. Date:	28-10-2023	Date of Issue:	30-10-2022	
DUC Receipt Date:	28-10-2022	SRF No.	221028.8	SRF Date:	28-10-2022	
Issued to (Customer Details):-	M/s.SDH Pattamundai Kendrapara,Odisha- 754215		Chalan/ Reference : ---			
Equipment Location:	Blood Bank		Chalan/ Reference Date :- ---			
Details of Equipment /Device Under Calibration(DUC):-						
Name of equipment:	Micropipette	Equipment ID:	N/A			
Make:	Microlit	Range/ Size:	5-50 μ L			
Model/ Type:	---	Least Count:	1 μ L			
Serial No.:	09034158	Accuracy / Tolerance:	Not provide by Customer			
KTPL Bar Code:	241191	Operational Manual:	Not provide by Customer			
Condition of DUC:	Satisfactory	Calibrated at (Lab/Site):	Lab			
Environmental Conditions During Calibration:						
Temperature in ($^{\circ}$ C):-	20 \pm 2 $^{\circ}$ C		Relative Humidity in (%):-	50 \pm 10 %		
Calibration Reference standard :						
KTPL/CP45-01-00		ISO 8655-1:2002 part 2 and euramate cg-19				
Detail of Reference standard Instruments used for Calibration:						
Description	Make/Model.	Serial No.	Certificate No.	Cal. Validity		
Digital Analytical Balance	A & D/BM-20	T1003047	ITCCS/2021/50936	17-Nov-2022		
Visual Inspection of Device Under Calibration :- OK						
Calibration/Performance result at 20$^{\circ}$C:-						
Sr. No.	Parameter	Unit of Measurement	DUC Value	STD Value	Error/ Deviation	Expended Uncertainty of measurement \pm(μL)
1)	Volume	(μ L)	50	49.60	0.40	0.6

Sarfaraz Ahmad

Calibrated By : (Sarfaraz Ahmad)



Santosh

Approved and Released By: (Santosh Kumar).

Calibration Certificate

Page 2 of 2

Certificate Number:	221028.8.2	Date of Issue:	30-10-2022
Calibration Date:	28-10-2022	Next Recommended Cal. Date:	28-10-2023
Note:			
1) The Equipment used for calibration of DUC are calibrated and traceable to national / international standards.			
2) This Certificate refers to only particular item (s) submitted for calibration.			
3) Kirloskar technologies (p) ltd. is not liable for any changes in calibration data and performance specification on account of malfunctioning of standards/ instruments/equipment covered by this certificate due to damaged caused to it after issuance of this certificate			
4) The calibration result reported in this certificate are valid at the time of and under stated conditions of measurement.			
5) This certificate shall not be reproduced in full/ part without prior permission of kirloskar technologies (p) ltd.			
6) All Precaution have been taken for any error or omission while calibrating the equipment and issuing its certificate. However Kirloskar Technologies (p) Ltd shall not be liable for any loss or liability that may be arise to any party in this regards.			
7) Equipment is calibrated at points as per customer request. Decision rule: Acceptance where "Pass" indicates measured value is within tolerance limit of accuracy.			
8) " Fail" indicates measured values are not within tolerance limit of accuracy without considering measurement uncertainty and is applicable only if Tolerance limits are provided by the customer or mentioned in Manufacture 'specification .			
Conclusion/Remarks:			
1) Operational manual of Not available Tolerances not provided by the customer:- Decision rule not applicable:			
2) The Reported Expanded uncertainty of measurement is calculated at approximately 95% confidence level with coverage factor k=2			

End of report

Sarfaraz Ahmad

Calibrated By : (Sarfaraz Ahmad)



Santosh

Approved and Released By: (Santosh Kumar).