

CALIBRATION CERTIFICATE

This is to Certify that Biochemistry Analyzer Meril **CliniQuant** Instrument Sr. No. **CQ110850** Installed at **Oxford Hospital** has been calibrated by our engineer on **06-Mar-2023** with appropriate and sophisticated tools. Below parameters are been calibrated:

TEST ITEMS AND RESULT

Appearance	OK
AD Value Range (120<AD<350)	OK
Temperature Checking (at 37'C)	OK
Printer Test	OK
Aspiration Volume Range	OK
Stability of Pot. Dichromate	OK
Absorbency Linearity at 340	OK
PC Communication Checking	OK
Sample Test 1	OK
Sample Test 2	OK
Sample Test 3	OK

This Calibration Certificate is Valid Up to **05-Mar-2024**.

Conclusion: All parameter are within range. The Machine is working properly. (As per Meril terms.)

Engineer Name: **Mr. Ajay Kumar**

Engineer Sign:



Test Result of biochemistry parameter before & after calibration

Parameter	Result before	Result after
Glucose	110	102
Bilirubin	1.0	0.9
Total protein	7.0	6.9
s. cretnine	1.0	0.9
Urea	29	28
Uric acid	3.4	3.6
Cholesterol	176	169
TG	130	126
Albumin	3.9	3.8

This Calibration Certificate is Valid Up to **05-Mar-2024**.

Conclusion: All parameter are within range. The Machine is working properly. (As per Meril terms.)

Engineer Name: **Mr. Ajay Kumar**

Engineer Sign:



T +91 260 3052100
F +91 260 3052125

www.merillife.com

Meril Diagnostics Pvt. Ltd.
Survey No. 135/2/B, Bilakhia House
Muktanand Marg, Chala
Vapi - 396191, Gujarat, India.

Cardiovascular

Orthopedic

Surgical

Diagnostics