

Certificate of Calibration

This is to certify that Fully Automatic Hematology Analyzer Model "Cell-Tech 380" having Sr. No. 665000215 installed at [OXFORD HOSPITAL, 305, LAJPAT NAGAR, NEAR, NAKODER CHOWK, JALANDHAR-144001] has been calibrated on 27-03-2023 by the technical staff of Arora Diagnostic.

This Calibration certificate is valid up to 26-09-2023.

For Arora Diagnostic

Shekhar
Arora Diagnostic
24, JAIN MARKET,
DR. NOHRIA STREET,
BATHINDA - 151001

Test Result of Hematology parameter before & after calibration

Parameter	Result before	Result after
HB	5.0	4.9
TLC	6300	6100
DLC	45,30,15,10	43,33,13,11
RBC	1.65	1.60
PCV	13.9	13.1
MCV	84.0	83.1
MCH	30.1	31.1
MCHC	35.8	34.9
PLATLETS	60000	59000