



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JUNE 2023



PC-1024

Lab Name VIJAY CLINICAL LAB Lab No 5938  
 Constituent Group Urine (Microalbumin & Protein) Date of Result Entered : 09/06/2023  
 PT item Lyophilized human urine based Date of Report Published : 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	UR. PROTEIN	Ur. Benzethonium	Any Analyser (Automation / Semi Automation)	4	15.79	4.83	0.76	15 mg/dL	-1.04	0.76
2	UR. MICROALBUMIN	Ur. Immuno Turbidimetric	Any Analyser (Automation / Semi Automation)	144	79.49	19.49	15.50	42 mg/L	-2.42	2.58

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

LAB ADDRESS :  
 VIJAY CLINICAL LAB  
 VIJAY DIABETES, THYROID & ENDOCRINE CLINIC, 12 FIRST CROSS, SARADAMBAL NAGAR  
 PUDUCHERRY  
 PUDUCHERI605005

Coordinator Contact Details:  
 Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
 Contact Number: 0416-2283102

*Pamela Christudoss*  
 Dr. Pamela Christudoss  
 CMC EQAS Coordinator  
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
 Data in CMC EQAS reports is confidential  
 CMC EQAS does not sub contract any components  
 \*\*\*\*\* End of Report \*\*\*\*\*