

UNIQUE CALIBRATION & CONSULTANT SERVICES

AN ISO 9001:2015 CERTIFIED CALIBRATION LABORATORY



Calibration Certificate

Format No.: UCCS/QF/7.8-09		Page No :- 1 of 1
Certificate No.	UCCS/21/A/167-013	
Issue Date	11.03.2023	
Service Request No.	UCCS//21/H/170	Calibration Certificate of
Service Request Date	11.03.2023	
Receipt Date	11.03.2023	
Calibration Date	11.03.2023	
Calibration Due Date (As per	10.03.2024	Micro pipette
Customer request)		

Name & Address of Customer	Identification on UUC (Unit Under Calibration)		
Swadha Hospital Kachora Road, Rail Mandi, Jaswanth Nagar, Etawah- 206245	Range/ Size Make/Model Instrument I.D. No. Instrument Sr. No. Least Count Location/Dept. Visual Inspection	5 to50 ul Perfect Xtra/REF-T/ML/15 0.1 ul MAIN LAB Ok	

Calibration Procedure	UCCS/TH/CP-04		Condition of	of Instrument	OK
Reference Standard	DKD R5-1		Calibration	Performed at	Lab
Environmental Condition	Temperature	30 :	±5°C	Relative Humidit	30 t0 80 % RH

Standard Equipment Used for Calibration					
Instrument Name	Name Make I.D No./Sr.No. Calibrated By		Certificate No.	Due Date	
Tycon Engineeing	TIPL	UCCL/ET/PC-01	Acumen Measurement Delhi	AMC/00001537	6/30/2024

Serial No	Nominal Value In (kg/cm2)	Observed Value In kg/cm2	Error in °C Mean	Uncertainty of Measurement ± °C
1	4.0	3.9	0.1	
2	8.0	7.8	0.2	± 0.46 °C
3	25.0	15.9	9.1	
4	50.0	20.0	30.0	

The reported uncertainty stated is the expanded uncertainty of measurement obtained

by multiplying the standard uncertainty by the coverage factor k=2 correspond to confidence level of 95 %

Note :- 1. This Certificate refers only to the particular items submitted for calibration.

 This Certificate shall not be reproduced, except in full, without the written permission of Managing director Calibration System, Rudrapur.

3. Result Reported are Valid at the of and under the stated conditions of measurement.

4. Laboratory Standards are traceable to National Standards.

5. Calibration Certificate issue for weight & Measure parameters i.e. Mass, Balance, Volumetric equipment,

Measuring Scale/Tapes etc.for scientific purpose only and should not be used for Trade / Commerical use.

End of Certificate

m Calibrated By Amit Kumar **Calibration Engineer**



Working : W.Z. -1390/Z-2, 1st Floor, Nangal Raya Extn., Pankha Road, New Delhi-110046 Regd. Office : R.Z.-56A, Indra Park, Gali No.25, Palam Colony, New Delhi-110045