



**Metropolis Healthcare Ltd., EQAS**  
MEDICENTRE SONOGRAPHY AND CLINICAL LAB, CHITTORGARH (454)

**Outlier And Analyte Summary Report**

Outlier Details For Cycle No 220412 and Sample No 01

Report Date : 25/01/2023

Analyte	Instrument	Result Value	Standard Unit
# Urine Pregnancy Test	-	Negative	-
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

**Problem Classification:** \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**Corrective Action:** \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**Reviewed by:** \_\_\_\_\_

**Dated:** \_\_\_\_\_

Instrument : Manual

Analyte	Standard Unit	Result Value	Accepted Value
@ pH	-	6	6,6.5,7
@ Specific Gravity	-	1.015	1.010,1.015,1.020
@ Glucose (urine)	-	2+	+2,+3,+4
@ Protein (Urine)	-	Negative	+1,Neg,Tr
@ Ketone (urine)	-	Negative	Neg,Tr
@ Blood (urine)	-	2+	+1,+2,Tr
@ Bilirubin (urine)	-	Negative	+1,Neg
@ Urobilinogen (urine)	-	Normal	Nor
@ Nitrite (urine)	-	Absent	Abs

Instrument : -

Analyte	Standard Unit	Result Value	Accepted Value
# Urine Pregnancy Test	-	Negative	
Legend @ : Acceptable		Total Parameters	10
@ : Acceptable		Not Evaluated Parameters	1
X : Unacceptable		Evaluated Parameters	9
# : Not Evaluated		Outlier Parameters (X)	0
⊙ : Delayed Result Entry		EQAS Score Clinical Path - Urine Routine Chemistry	100.00 %



**Dr Puneet Kumar Nigam**  
PT coordinator & Technical Manager, MHL EQAS  
Unit No. 409-416.  
Commercial Building - 1A  
Kohinoor Mall, Kiroil Road, Kurla (W),  
Mumbai - 400070