

Outlier And Analyte Summary Report

Outlier Details For Cycle No 220309 and Sample No 03

Report Date : 10/12/2022

Analyte	Instrument	Result Value	Standard Unit	Z-Score
! MCHC (Mean Corpuscular Hb Concentration)	Sysmex XS 800 I	20.20	g/dL	-2.38
Legend @ : z score ≤ 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score ≥ 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Sysmex XS 800 I

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	1.33	1.30	1.12	1.19
@ Total Leucocytes (WBC) count	/c.mm	1570.00	1540.00	0.16	-0.34
@ Haemoglobin	g/dL	3.40	3.35	0.24	0.61
@ MCV (Mean Corpuscular Volume)	fL	120.60	121.00	0.89	0.66
@ MCH (Mean Corpuscular Hb)	pg	25.60	25.90	-0.70	-0.45
! MCHC (Mean Corpuscular Hb Concentration)	g/dL	20.20	20.90	-2.38	-1.36
@ Platelet Count	1000/ µL	38.00	34.50	1.17	-0.11

Legend @ : Acceptable
! : 2.0 < z score < 3.0 - Warning
X : z score ≥3 - Unacceptable
: Not Evaluated
⊖ : Delayed Result Entry
* : Not considered for evaluation.

Total Parameters	7
Not Evaluated Parameters	0
Evaluated Parameters	7
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kirol Road, Kurla (W),
Mumbai - 400070