



Alere Medical Private Limited
No. 404, BPTP Park Centra,
NH - 8, Sector - 30
Opp. 32nd Milestone, Gurugram
Haryana 122001, India
T: +91 124 4569000
F: +91 124 4569011

CALIBRATION REPORT- UROMETER-120

CUSTOMER NAME	MEDICENTRE SONOGRAPHY & LAB	
ADDRESS	15, MAHADEV COLONY, NEAR M.G. HOSPITAL, BANSWARA	
NAME OF THE INSTRUMENT	URINE ANALYZER	
MODEL	UROMETER-120	
INSTRUMENT S/N	1275029	
DATE OF CALIBRATION	30.01.2023	
DUE DATE OF CALIBRATION	29.01.2024	
CALIBRATION MATERIAL	RANGE	RESULT
UROMETER WHITE STRIP	333 +/- 10	WITHIN RANGE
Q.C. MATERIAL FOR CALIBRATION CHECK	RESULT	REMARK
UROMETER LOW	WITHIN RANGE	PASS
UROMETER HIGH	WITHIN RANGE	PASS

For Alere Medical Pvt. Ltd.



Authorized Signatory