

Date : 08/02/2023

CALIBRATION CERTIFICATE

AA Diagnostics Center
Rajans Hospital
Ramanathapuram

Dear Sir/ Madam ,

As desired by you please find the enclosed herewith the calibration report.

1. Serial No : 60370
2. Range of the instrument : Optical density Liner from 0-3.5 A
3. Precision & Accuracy study on EM 360 done on 08/02/2023.

Next Calibration due on or before 08/02/2024

This instrument is manufactured in accordance with the European law in force, in particular to “Electricity Safety” EN 610101 and Electromagnetic Compatibility “ EN 61326 Standards.

WAVE LENGTH	GAIN
340	120
376	126
415	137
450	139
480	123
505	129
546	131
570	139
600	141
660	143
700	140
750	132

TRANSASIA BIO-MEDICALS LTD.

Transasia House, 8 Chandivali Studio Road, Andheri (E), Mumbai - 400 072 Tel.: (022) 40309000, Fax: (022) 2857 3030
Email : transasia@transasia.co.in Website : www.transasia.co.in


	IDEAL	ACTUAL
Pressure	0.99 bar	1.0 bar
Lamp Voltage	12 V	11.74V
Temperature	37 C	37 C

Photometer Calibration Check	OK
Cuvette Blank Check	OK

The instrument does photometer check before each batch of tests is run. It will give an error message if the photometer test fails and it will not run any samples.

Thanking you and assuring of our prompt attention at all times.

Your faithfully,
For Transasia-Bio Medicals Ltd,


R.BALAMURUGAN
Sr.Service Engineer

Date: 09/02/2023

CALIBRATION CERTIFICATE**TO WHOM SOEVER IT MAY CONCERN**

This is to certify that the Erba Chem7 Biochemistry analyzer **Serial No: N140277** has installed by us at **AA DIAGNOSTIC CENTER**, Rajans hospital, Ramanathapuram has been calibrated on 09-Feb-2023 and verified to control values.

The Next Calibration due is on or before : **09-Feb-2024**

For TRANSASIA BIO MEDICALS LTD



R. BALAMURUGAN
SR.SERVICE ENGINEER

Date: 02/03/2023

CALIBRATION CERTIFICATE**TO WHOM SOEVER IT MAY CONCERN**

This is to certify that the Easylyte analyzer bearing the serial no: 65638CNKC has installed by us at AA DIAGNOSTIC CENTRE, Ramanathapuram is automatic self calibrating electrolyte analyser for estimation of Sodium, Pottasium, Chloride.

The following Diagnostics checks were Carried in Month of 01-03-2023

The Next Calibration due is on or before : 01-Mar-2024

Sodium Calibration Value. - 58.94
Pottasium Calibration Value - 57.57
Chloride Calibration Value - 48.04

For TRANSASIA BIO MEDICALS LTD



R. BALAMURUGAN
SR.SERVICE ENGINEER

TRANSASIA BIO-MEDICALS LTD.

Transasia House, 8 Chandivali Studio Road, Andheri (E), Mumbai - 400 072 Tel.: (022) 40309000, Fax : (022) 2857 3030
Email : transasia@transasia.co.in Website : www.transasia.co.in