



Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments
 304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092
 Phone / Fax : (011) 22453259, 22434362, 9599482307
 E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

CALIBRATION CERTIFICATE

Page 1 of 1

CALIBRATION CERTIFICATE OF DIGITAL TEMPERATURE INDICATOR WITH SENSOR

Certificate No:	23100009859	Issued On:	29/05/2023
ULR No:	CC214623100009859F	Calibration Date:	26/05/2023
Job Identification No:	ICS/C/T/05/9859	Next Calibration Date:	25/05/2024
Ref. No:	SRF, Dated - 22/05/2023	Receipt Date:	24/05/2023

CALIBRATED FOR: Dr. Pankaj Diagnostic & Research Centre Pvt. Ltd.
 B/27, Sachiwalaya Colony,
 Kankarbagh, Patna-800020.

EQUIPMENT DESCRIPTION

Name	Digital Temperature Indicator with Sensor		
Make/Model	Multi / ST-9283B	Visual Inspection	Ok
Serial No.	---	I.D. No.	---
Range	-50°C to 300°C	Least Count	0.1° C
Location	---	Calibration Site	In-Lab

ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 3.0° C	Humidity	50 ± 10 % RH
-------------	------------------	----------	--------------

STANDARD REFERENCE DETAILS Traceable to National / International Standard

Name of the Standard	Make	Certificate No.	Calibration Agency/Traceability	Date of Calibration	Valid up to
RTD Sensor with Digital Indicator (S-1)	Yudian	TL/023/317.1.1	Tempsens, Udaipur	11-03-2023	10-03-2024
RTD Sensor with Digital Indicator (S-2)	Yudian	TL/023/317.2.1	Tempsens, Udaipur	11-03-2023	10-03-2024

USED EQUIPMENT DETAILS	Methanol Bath, Water Bath & Dry Block Calibrator.
PRINCIPLE/METHODOLOGY OF CALIBRATION:	As per Calibration Procedure No.: ICS/CAL/SOP-TH02 (ITS-90) (Comparison Method)

RESULTS: Thermal Calibration (Temperature)		
U.U.C. READING (In °C)	CORRECTED VALUE OF STANDARD (in °C) (Avg.)	Uncertainty at approx 95% C.L and coverage factor k=2 (in °C)
-30.0	-30.208	± 0.11
50.0	49.752	± 0.11
100.0	99.647	± 0.25
200.0	199.523	± 0.42
300.0	299.401	± 0.21

REMARKS: The recommended date for next calibration is mentioned, as asked by the customer.

END OF REPORT

- Note:**
- This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
 - The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
 - Total liability of our Organisation is limited to the invoiced amount.
 - Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
 - In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
 - The calibration certificate/Test Report is valid only for the condition of the UUC at the time under stated condition of calibration.

For and on behalf of
Indian Calibration Services


 Checked by


 RAJESH KAPOOR
 Operation-Head