

DAM

# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - APRIL 2023



PC-1024

Lab Name: Govt. PHC, Kadamangudi

Lab No: 9238

Department: Chemistry III

Date of Result Entered: 19/04/2023

Specimen: Lyophilized human serum based

Date of Report Published: 06/05/2023

A	Name of Principle	Analyzer Name	No. Participants	DV	Participants		Your Value	SDI	U
					CV	SD			
GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	4832	119.45	6.04	7.21	119 mg/dL	-0.06	0.34
UREA III	Urease UV / GLDH III	Robonik ( Semi automatic / Fully Automatic )	169	42.32	10.68	4.52	35 mg/dL	-1.62	0.70
URIC ACID III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1820	2.16	8.09	0.18	2.1 mg/dL	-0.34	0.01
CHOLESTEROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1750	111.61	5.19	5.79	108 mg/dL	-0.62	0.28

SDI Range	Interpretation
0 to 1.00	Excellent.

*[Signature]*  
**MEDICAL OFFICER**  
 Govt. Primary Health Centre  
 Kadamangudi  
 Kodavasal (TN)