

CALIBRATION REPORT

Centrifuge Table top



Report No:	CAL-AR-2023-02-0541	Date:09-01-2023	Page 01 of 02
------------	---------------------	-----------------	---------------

Customer Details			
Health Facility	Namsai DH		
State	Arunachal Pradesh	District	Namsai

Medical Equipment Details			
ME Number	ARZ1NAMDH0190173	Manufacturer	Unilab
ME Name	Centrifuges, Tabletop (Centrifuge)	Model	RH 142
Department	ICTC, Near ECG Room	Serial No	NA
Calibrated at	On Site	Condition of ME	Good
Date of Calibration	02-01-2023	Due for Calibration	01-01-2024

Environmental Conditions			
Temperature in °C	20°C	Humidity in % Rh	58%



Details of Master Equipment Used for Calibration						
No	Equipment ID	Description	Manufacturer	Model	Serial No.	Calibration Validity
1	TRI-TE-0015	DIGITAL TACHOMETER	AMBROBE	TACH-10	16060021	15-11-2023

Performance Results						
No	Parameters	Unit	Set Value	MeasuredValue	Acceptable Range	Remarks
1	Speed	rpm	1500	1490	1425 to 1575	Pass
2	Speed	rpm	2500	2600	2375 to 2625	Pass
3	Speed	rpm	3500	3450	3325 to 3675	Pass

CALIBRATION REPORT

Centrifuge Table top



Conditions of Issue of report	
1. The Standards used are traceable to National standards.	
2. This report should not be reproduced except in full without written approval.	
Calibration Status	Passed <input checked="" type="checkbox"/> Failed <input type="checkbox"/>
Calibrated By  Tapash Mazumdar	Authorised Signatory  Rahul Kumar Kasyap