

CALIBRATION REPORT

Refrigerators, Laboratory(Lab refrigerator)



Report No:	CAL-AR-2023-02-0540	Date:12-01-2023	Page 01 of 02
------------	---------------------	-----------------	---------------

Customer Details			
Health Facility	Namsai DH		
State	Arunachal Pradesh	District	Namsai

Medical Equipment Details			
ME Number	ARZ1NAMDH0190172	Manufacturer	LG
ME Name	Refrigerators, Laboratory(Lab refrigerator)	Model	(175L)
Department	ICTC, Near ECG Room	Serial No	9G2R0411
Calibrated at	On Site	Condition of ME	Good
Date of Calibration	03-01-2023	Due for Calibration	02-01-2024

Environmental Conditions			
Temperature in °C	20°C	Humidity in % Rh	58%

Details of Master Equipment Used for Calibration						
No	Equipment ID	Description	Manufacturer	Model	Serial No.	Calibration Validity
1	TRI-TE-0187	DIGITAL THERMOMETER	METRAVI	DTM-900	271926/1	15-11-2023

Performance Results						
No	Parameter	Set Temp	Unit	Measured Value	Acceptable Range	Remarks
1	Temperature	2°C	°C	2.1	1.5 to 2.5°C	Pass
2	Temperature	4°C	°C	4.3	3.5 to 4.5°C	Pass

CALIBRATION REPORT 1

Refrigerators, Laboratory(Lab refrigerator)



3	Temperature	6°C	°C	6.2	5.5 to 6.5°C	Pass
---	-------------	-----	----	-----	--------------	------

Conditions of Issue of report

1. The Standards used are traceable to National standards.
2. This report should not be reproduced except in full without written approval from Trimed Solutions India Private Limited.

Calibration Status

Passed

Failed

Calibrated By

Handwritten signature of Surojit Dey in black ink.

Surojit Dey

Authorised Signatory

Handwritten signature of Rahul Kumar Kashyap in black ink.

Rahul Kumar Kashyap