



Alere Medical Pvt. Ltd.
No. 404, BPTP Park Centra,
NH - 8, Sector - 30
Opp. 32nd Milestone, Gurugram
Haryana 122001, India
T: +91 124 4569000
F: +91 124 4569011

Calibration Certificate

Instrument Name : Urometer120
Serial Number : 12211058
Customer Name : Life Well Diagnostics (P) Ltd.
Address : LifeWell Diagnostics (P) Ltd, 1st Floor , Gurudwara Buildg,
Shri GuruSingh Sabha , Sector 19, Chandigarh-160019. Punjab.
Standard used : Self Calibration Mode
Standard Traceability : In House Certificate

Calibration

The instrument function well in all the systems check, self check and maintenance Procedures. The equipment was calibration and the Red, Blue & Green (R,B & G) filters of all the parameters and top & bottom were found to be within the acceptable range of 333 ± 10 .

QC Performance : Acceptable
Self Calibration : Acceptable

Performance Qualification : Passed

Calibration Date : 13.Apr.2023
Calibration : Passed
Calibration Valid up to : 12.Apr.2024

For Alere Medical Pvt Ltd.

(R. Muralidharan)
Engineer Technical Services-Kerala
Mob:-09567271430



REGISTERED OFFICE: 221 & 222, DLF TOWERS, SHIVAJI MARG, NAJAFGARH ROAD, NEW DELHI - 110015 (INDIA)
TEL: +91-11-45089400
CIN NO: - U74899DL1995PTC069456

alere.com



Positive CC

APR/13/2023 10:59:07

Name:
Ward:

SD (10) (01201)

Sex:
Age:

BLD	++	50RBC/ul
BIL	+++	3.0mg/dl
URO	+	1.0mg/dl
KET	+	10mg/dl
PRO	++	100mg/dl
NIT	pos	0.1mg/dl
GLU	+++	1000mg/dl
pH		6.0
S.G		≥1.025
LEU	+	25WBC/ul

(SN=00000000)

ID()

OP(00000000)LOT(00000000)

Positive CC

APR/13/2023 10:59:18

Name:

Ward:

SD (10) (01202)

Sex:
Age:

BLD	++	50RBC/ul
BIL	+++	3.0mg/dl
URO	+	1.0mg/dl
KET	+	10mg/dl
PRO	++	100mg/dl
NIT	pos	0.1mg/dl
GLU	+++	1000mg/dl
pH		6.0
S.G		≥1.025
LEU	+	25WBC/ul

(SN=00000000)

ID()

OP(00000000)LOT(00000000)

Positive CC

APR/13/2023 10:59:30

Name:

Ward:

SD (10) (01203)

Sex:
Age:

BLD	++	50RBC/ul
BIL	+++	3.0mg/dl
URO	+	1.0mg/dl
KET	+	10mg/dl
PRO	++	100mg/dl
NIT	pos	0.1mg/dl
GLU	+++	1000mg/dl
pH		6.0
S.G		≥1.025
LEU	+	25WBC/ul


For Alere Medical Pvt Ltd

(Signature)

(R. Muralidharan)

Engineer Technical Services-Kerala

Mob:-09567270430 LOT(00000000)



Alere Medical Pvt. Ltd.

No. 404, BPTP Park Centra,

NH - 8, Sector - 30

Opp. 32nd Milestone, Gurugram

Haryana 122001, India

T: +91 124 4569000

Sex: +91 124 4569011

Positive CC

Calibration Certificate

APR/13/2023 10:59:41

Name:

SD (10) (01204)

Sex:

Age:

BLD	++	50RBC/ul
BIL	+++	3.0mg/dl
URO	+	1.0mg/dl
KET	+	10mg/dl
PRO	++	100mg/dl
NIT	pos	0.1mg/dl
GLU	+++	1000mg/dl
pH		6.0
S.G		≥1.025
LEU	+	25WBC/ul

(SN=00000000)

ID()

OP(00000000)LOT(00000000)

Positive CC

APR/13/2023 11:00:06

Name:

Ward:

SD (10) (01206)

Sex:
Age:

BLD	++	50RBC/ul
BIL	+++	3.0mg/dl
URO	+	1.0mg/dl
PRO	++	100mg/dl
NIT	pos	0.1mg/dl
GLU	+++	1000mg/dl
pH		6.0
S.G		≥1.025
LEU	+	25WBC/ul

(SN=00000000)

ID()

OP(00000000)LOT(00000000)



CALIBRATION REPORT
APR/13/2023 10:33:54

** Calibration status **

Abbott

BLD	R=334	G=332	B=334
BIL	R=334	G=334	B=332
URO	R=334	G=332	B=334
KET	R=335	G=331	B=334
PRO	R=333	G=333	B=334
NIT	R=334	G=332	B=334
GLU	R=334	G=333	B=333
pH	R=335	G=332	B=333
S.G	R=334	G=332	B=334
LEU	R=335	G=332	B=333
VTC	R=335	G=331	B=334
REF	R=335	G=331	B=334
TOP	R=335	G=334	B=334
BOT	R=333	G=331	B=332

(SN=00000000)

Negative QC

APR/13/2023 11:31:56
Name:
Ward:
SD (10) (01229)

Sex:
Age:

BLD	-	neg
BIL	-	neg
URO	norm	0.1mg/dl
KET	-	neg
PRO	-	neg
NIT	-	neg
GLU	-	neg
pH	-	6.0
S.G	-	1.020
LEU	-	neg

(SN=00000000)
ID()
OP(00000000)LOT(00000000)

Negative QC

APR/13/2023 11:32:41

Name:
Ward:
SD (10) (01232)

Sex:
Age:

BLD	-	neg
BIL	-	neg
URO	norm	0.1mg/dl
KET	-	neg
PRO	-	neg
NIT	-	neg
GLU	-	neg
pH	-	6.0
S.G	-	1.020
LEU	-	neg

(SN=00000000)
ID()
OP(00000000)LOT(00000000)

Negative QC

APR/13/2023 11:34:06

Name:
Ward:
SD (10) (01234)

Sex:
Age:

BLD	-	neg
BIL	-	neg
URO	norm	0.1mg/dl
KET	-	neg
PRO	-	neg
NIT	-	neg
GLU	-	neg
pH	-	6.5
S.G	-	1.025
LEU	-	neg

(SN=00000000)
ID()
OP(00000000)LOT(00000000)

For Alere Medical Pvt Ltd

(R. Muralidharan)
Engineer Technical Services-Kerala
Mob:-09567271430

(SN=00000000)
ID()
OP(00000000)LOT(00000000)

REGISTERED OFFICE: 221 & 222, DLF TOWERS, SHIVAJI MARG, NAJAFGARH ROAD, NEW DELHI - 110015 (INDIA)

TEL: +91-11-45089400
CIN NO: - U74899DL1995PTC069456

Alere Medical Pvt. Ltd.
No. 404, BPTP Park Centra,
NH - 8, Sector - 30
Opp. 32nd Milestone, Gurugram
Haryana 122001, India

T: +91 124 4569000
F: +91 124 4569011

Calibration Certificate

BLD	R=332	G=332	B=338
BIL	R=332	G=332	B=336
URO	R=332	G=332	B=336
KET	R=332	G=333	B=335
PRO	R=331	G=334	B=335
NIT	R=332	G=332	B=336
GLU	R=334	G=331	B=336
pH	R=333	G=333	B=336
S.G	R=332	G=332	B=336
LEU	R=334	G=331	B=336
VTC	R=332	G=332	B=336
REF	R=334	G=331	B=335
TOP	R=334	G=334	B=338
BOT	R=331	G=330	B=334

(SN=00000000)

Negative QC

APR/13/2023 11:34:45

Name:
Ward:
SD (10) (01235)

Sex:
Age:

BLD	-	neg
BIL	-	neg
URO	norm	0.1mg/dl
KET	-	neg
PRO	-	neg
NIT	-	neg
GLU	-	neg
pH	-	6.0
S.G	-	1.020
LEU	-	neg

(SN=00000000)
ID()
OP(00000000)LOT(00000000)

