



Alere Medical Pvt. Ltd.
No. 404, BPTP Park Centre,
NH- 8, Sector- 30
Opp: 32nd Milestone, Gurugram
Haryana 122061, India
T: +91 124 4589000
F: +91 124 4589011

Calibration Certificate

Instrument Name : Urometer120
Serial Number : 12211058
Customer Name : Life Well Diagnostics (P) Ltd.
Address : LifeWell Diagnostics (P) Ltd, 1st Floor , Gurudwara Buildg,
Shri GuruSingh Sabha , Sector 19, Chandigarh-160019. Punjab.
Standard used : Self Calibration Mode
Standard Traceability : In House Certificate

Calibration

The instrument function well in all the systems check, self check and maintenance Procedures. The equipment was calibration and the Red, Blue & Green (R,B & G) filters of all the parameters and top & bottom were found to be within the acceptable range of 333_{+10} .

QC Performance : Acceptable

Self Calibration : Acceptable

Performance Qualification : Passed

Calibration Date : 13.Apr.2023

Calibration : **Passed**

Calibration Valid up to : 12.Apr.2024

For Alere Medical Pvt Ltd.

(R. Muralidharan)
Engineer Technical Services-Kerala
Mob:-09567271430





Positive CC

APR/13/2023 10:59:07

Name:
Ward:

Sex:
Age:

Abbott

SD (10) (01201)

BLD	++	50RBC/ul
BIL	+++	3.0mg/dl
URO	+	1.0mg/dl
KET	+	10mg/dl
PRO	++	100mg/dl
NIT	pos	0.1mg/dl
GLU	+++	1000mg/dl
pH		6.0
S.G		≥1.025
LEU	+	25WBC/ul

(SN=00000000)

ID()

OP(00000000)LOT(00000000)

Positive CC

APR/13/2023 10:59:18

Name:
Ward:

Sex:
Age:

SD (10) (01202)

BLD	++	50RBC/ul
BIL	+++	3.0mg/dl
URO	+	1.0mg/dl
KET	+	10mg/dl
PRO	++	100mg/dl
NIT	pos	0.1mg/dl
GLU	+++	1000mg/dl
pH		6.0
S.G		≥1.025
LEU	+	25WBC/ul

(SN=00000000)

ID()

OP(00000000)LOT(00000000)

Positive CC

APR/13/2023 10:59:30

Name:
Ward:

Sex:
Age:

SD (10) (01203)

BLD	++	50RBC/ul
BIL	+++	3.0mg/dl
URO	+	1.0mg/dl
KET	+	10mg/dl
PRO	++	100mg/dl
NIT	pos	0.1mg/dl
GLU	+++	1000mg/dl
pH		6.0
S.G		≥1.025
LEU	+	25WBC/ul

For Alere Medical Pvt Ltd

(R. Muralidharan)

Engineer Technical Services-Kerala

Mob:-09567270430



Alere Medical Pvt. Ltd.
No. 404, BPM Park Centre,
NH - 8, Sector - 30
Opp. 32nd Milestone, Gurugram
Haryana 122001, India.

Positive CC

APR/13/2023 10:59:41

Name:
Ward:

Sex:
Age:

Calibration Certificate

SD (10) (01204)

BLD	++	50RBC/ul
BIL	+++	3.0mg/dl
URO	+	1.0mg/dl
KET	+	10mg/dl
PRO	++	100mg/dl
NIT	pos	0.1mg/dl
GLU	+++	1000mg/dl
pH		6.0
S.G		≥1.025
LEU	+	25WBC/ul

(SN=00000000)

ID()

OP(00000000)LOT(00000000)

Positive CC

APR/13/2023 11:00:06

Name:
Ward:

Sex:
Age:

SD (10) (01206)

BLD	++	50RBC/ul
BIL	+++	3.0mg/dl
URO	+	1.0mg/dl
KET	+	10mg/dl
PRO	++	100mg/dl
NIT	pos	0.1mg/dl
GLU	+++	1000mg/dl
pH		6.0
S.G		≥1.025
LEU	+	25WBC/ul

(SN=00000000)

ID()

OP(00000000)LOT(00000000)



Kulraj

REGISTERED OFFICE: 221 & 222, DLF TOWERS, SHIVAJI MARG, NAJAFGARH ROAD, NEW DELHI - 110015 (INDIA)

TEL: +91-11-45069400

CIN NO: - U74899DL1995PTC069456

alere.com

** Calibration status **
Abbott

** Calibration status **

BLD	R=334	G=332	B=334
BIL	R=334	G=334	B=332
URO	R=334	G=332	B=334
KET	R=335	G=331	B=334
PRO	R=333	G=333	B=334
NIT	R=334	G=332	B=334
GLU	R=334	G=333	B=333
pH	R=335	G=332	B=333
S.G	R=334	G=332	B=334
LEU	R=335	G=332	B=333
VTC	R=335	G=331	B=334
REF	R=335	G=331	B=334
TOP	R=335	G=334	B=334
BOT	R=333	G=331	B=332

Calibration Certificate

BLD	R=332	G=332	B=332
BIL	R=332	G=332	B=332
URO	R=332	G=332	B=332
KET	R=332	G=333	B=332
PRO	R=331	G=334	B=332
NIT	R=332	G=332	B=332
GLU	R=334	G=331	B=332
pH	R=333	G=333	B=332
S.G	R=332	G=332	B=332
LEU	R=334	G=331	B=332
VTC	R=332	G=332	B=332
REF	R=334	G=331	B=332
TOP	R=334	G=334	B=338
BOT	R=331	G=330	B=334

(SN=00000000)

(SN=00000000)

Negative QC

Negative QC

APR/13/2023 11:31:56
Name: Sex:
Ward: Age:
SD (10) (01229)

APR/13/2023 11:32:41
Name: Sex:
Ward: Age:
SD (10) (01232)

BLD	-	neg
BIL	-	neg
URO	norm	0.1mg/dl
KET	-	neg
PRO	-	neg
NIT	-	neg
GLU	-	neg
pH	-	6.0
S.G	-	1.020
LEU	-	neg

BLD	-	neg
BIL	-	neg
URO	norm	0.1mg/dl
KET	-	neg
PRO	-	neg
NIT	-	neg
GLU	-	neg
pH	-	6.0
S.G	-	1.020
LEU	-	neg

(SN=00000000)
ID()
OP(00000000)LOT(00000000)

(SN=00000000)
ID()
OP(00000000)LOT(00000000)

Negative QC

APR/13/2023 11:34:45
Name: Sex:
Ward: Age:
SD (10) (01235)

BLD	-	neg
BIL	-	neg
URO	norm	0.1mg/dl
KET	-	neg
PRO	-	neg
NIT	-	neg
GLU	-	neg
pH	-	6.0
S.G	-	1.020
LEU	-	neg

Negative QC

Negative QC

APR/13/2023 11:32:10
Name: Sex:
Ward: Age:
SD (10) (01230)

APR/13/2023 11:34:06
Name: Sex:
Ward: Age:
SD (10) (01234)

BLD	-	neg
BIL	-	neg
URO	norm	0.1mg/dl
KET	-	neg
PRO	-	neg
NIT	-	neg
GLU	-	neg
pH	-	6.0
S.G	-	1.020
LEU	-	neg

BLD	-	neg
BIL	-	neg
URO	norm	0.1mg/dl
KET	-	neg
PRO	-	neg
NIT	-	neg
GLU	-	neg
pH	-	6.5
S.G	-	1.025
LEU	-	neg

For Alere Medical Pvt Ltd

(R. Muralidharan)
Engineer Technical Services-Kerala
Mob:-09567271430

(SN=00000000)
ID()
OP(00000000)LOT(00000000)

(SN=00000000)
ID()
OP(00000000)LOT(00000000)



Handwritten signature

(SN=00000000)
ID()
OP(00000000)LOT(00000000)