

# OFFICE OF THE NODAL OFFICER

## ICTC-PPTCT Programme

Dept. of O&G, Silchar Medical College & Hospital, Silchar-14

Ref. No. SMC/O&G/AIDS/2004/-006

Date 15/03/2023

To  
The Incharge  
State Reference Laboratory.  
Silchar Medical College & Hospital.  
Silchar - 788014

Ref. NO: SMC/MICRO/SRL/05/23/614 Dated. Silchar the 25/02/2023

Sub: Submission of Panel Sera (1st round 2022-23) test report of PPTCT of as Part of EQAS Activity.

Sir/Madam

I have the honour to submit you the Panel Sera result of one set (04 nos) of (1st round 2022-23) EQAS Panel.

This is for favour of your kind information.

The result copy is attached in a prescribed format along with the letter.

With regards.

Your's faithfully  
15/3/23

(Dr. Alaka Banerjee)

Nodal Officer PPTCT  
Deptt. of O&G.  
SMCH, Silchar.

Received  
15/03/2023  
Ramu Kr. Das  
Lab. Technician, SRL  
SMCH Silchar

**NACO-SRL, SMCH SILCHAR**

External Quality Assessment Scheme for HIV Serology

(REPORTING FORMAT)

Name of ICTC / PPTCT Centre: Silchar Medical College & Hospital Result sheet of Panel Sera 1<sup>st</sup> round 2022-23.

Received date of Panel Sera: 27.02.2023 Reporting Date of Panel Sera: 15/03/2023

(Note: Column not to be blank)

Sample ID	Result	Details of Test kit used by - ICTC/PPTCT		
		Test - 1	Test - 2	Test - 3
		Kit Name: <u>CombaidS</u> Lot: <u>4000027449</u> Exp date: <u>2024.07.29</u> Manufacture: <u>2022.07.29</u> Principle of the test: <u>Dot immuno assay</u>	Kit Name: <u>StandardG</u> Lot: <u>2057073</u> Exp date: <u>05.2024</u> Manufacture: <u>06.2022</u> Principle of the test: <u>Immuno chromatographic assay</u>	Kit Name: <u>AIDSCAN</u> Lot: <u>37230104</u> Exp date: <u>06.2024</u> Manufacture: <u>01.2023</u> Principle of the test: <u>Immuno chromatographic assay</u>
SRL/SMCH/E//HIV -01	Negative	Non Reactive	NA	NA
SRL/SMCH/E//HIV -02	Negative	Non Reactive	NA	NA
SRL/SMCH/E//HIV -03	Positive	Reactive	Reactive HIV 1	Reactive HIV 1
SRL/SMCH/E//HIV -04	Negative	Non Reactive	NA	NA

Lab. Technician  
 Name & Signature: [Signature]  
 ICTC-PPTCT, SMCH  
 Silchar, Assam  
 Lab. Technician ICTC/PPTCT

11/03  
 Name & Signature: [Signature]  
 PPTCT Centre, SMCH  
 Silchar-14, Cachar  
 In-charge ICTC / PPTCT

R: Reactive; NR: Nonreactive; P: Positive; N: Negative; I: Indeterminate.

Remarks: (if any): 1st and 2nd test done on 28.02.23

Due to shortage of 3rd test kit, 3rd test done on 11.03.2023



NACO

OFFICE OF THE HIV TESTING LABORATORY (SRL, ICTC & CD4)

DEPARTMENT OF MICROBIOLOGY  
SILCHAR MEDICAL COLLEGE & HOSPITAL

SILCHAR-788014 (ASSAM)

(NABL ACCREDITED)

Email- srl.assam.smc@gmail.com

Telephone no – 03842240179



No. SMC/MICRO/SRL/05/23.....65).....Dated Silchar the 21/05/2023

To

The Medical Officer In-charge

PPTCT, Silchar Medical College &  
Hospital Silchar

Sub:- Proficiency Testing report ( 1<sup>st</sup> round, 2022-23)

Sir /Madam,

We are happy that your centre...PPTCT, Silchar Medical College & Hosp...has tested the Panel sera and sent the result to us. I would like to inform you that the centre has detected the HIV Panel sera correctly and is found to be Concordant.

Your centre has scored 100%

Looking towards future participation

With Regards

*[Signature]* 21/5/23

Dr. (Mrs) Debadatta Dhar Chanda

Professor & Head

Cum- Lab. Director & Quality Manager, HTL

Department of Microbiology

Silchar Medical College & Hospital Silchar

Lab. Director

HIV Testing Laboratory

SMCH Silchar

Received  
Lab. Technician  
ICTC-PPTCT, SMCH  
Silchar, Assam  
20/05/23