



Quality Certification & Laboratories

Plot No. 10, First Floor, D.S.I.D.C, Scheme-III,
Okhla Industrial Area, Phase- II, New Delhi - 110020, India
Tel: +91 9818009007
E-mail: corp@qcl.co.in

CALIBRATION CERTIFICATE

Certificate No. : QCL/DEL/CAL/2302105001.13
Date of Calibration: 10.02.2023
Date of Receipt of Instruments: 10.02.2023

Page 1 of 1
SRF No. : QCL/2302105001
Suggested Due Date: 10.02.2024
Date of Issue: 11.02.2023

Customer Name & Address:

M/s: CNS Hospital,
GATA NO 1982 BARABANKI-215122

Description of Instrument:

Name :	Automated Cell Counter	Range	As Per Range
Make :	Erba	Least Count	As Per Range
S. No. :	K10012220411	Accuracy	-----
Model No. :	H360	Performed at	Site
Location:	Pathology (New Building)	Visual Inspection	Ok
Customer ID:	CNS/ACC/01	Zero Error	Nil

Environmental Condition:

Temperature : 25°C ± 5°C Humidity : 45 % to 75%

Methodology of Calibration:

Reference Standard : IEC-60601/62353, IEC-60601-2-20 Work Instruction No. : QCL/WI/MED/17

Standard & Major Equipment (s) Used for calibration:

S. No	Instrument Name	Make	S. No./Model No.	Certificate No.	Calibrated By	Valid Upto
1	Electrical Safety Analyzer/ESA 609	Fluke	QCL/DEL/MED/014, 5188045	TSC/21-22/17900-3	Transcal	15.03.2023

Calibration Result:

Electrical Safety Test as per IEC-60601

Sr. No.	Parameter	Unit Of Measurement	Measured Value	Uncertainty of Measurement
1	Protective Earth Resistance	Ω	0.147	± 7.72%
2	Earth Leakage Current (NC)	μA	104.8	± 1.81%
3	Earth Leakage Current (SFC)	μA	193	± 1.88%
4	Enclosure Leakage Current (NC)	μA	2.1	± 5.86%
5	Enclosure Leakage Current (SFC)	μA	10.3	± 1.69%
8	Insulation Resistance (MΩ)	MΩ	61.1	± 8.29%

(The reported expanded Uncertainty in measurement is stated as the standard uncertainty in measurement multiplied by the coverage factor k = 2, which for a normal distribution corresponds to a coverage probability of approximately 95%)

Note:

- *The calibration results reported in this calibration certificate are valid at the time of & under stated conditions of measurement
- *This certificate cannot be reproduced except in full without our prior permission in writing
- *This certificate refers only to the particular item (s) calibrated.
- *Laboratory standard are traceable to national standards
- *DUC: Device Under Calibration
- *Is not in our NABL Scope Of Accreditation

Prepared By

Approved By

Customer Service Cell (Sheikh Salman)

Technical Manager (Rajesh Kumar)