

Instrument Calibration

Customer Name : Dr. Gunnas Diagnostics.
Customer Address : House No: 4-2, Street No-7, Beside NGRI
Metro station, Habsiguda,
Hyderabad-500007.
Contact Person : SM MUSTAFA.
Instrument Model : BC-720R-AUTO HEMATOLOGY ANALYZER.
Serial No : DE3-25000034.
Lot No : PLUS 1122 **Exp. Date** : 12-05-2023
Process Date : 27-01-2023

Parameters	WBC	RBC	HGB	MCV	PLT
Target	9.30	4.36	15.0	101.5	253
1	9.40	4.35	15.0	101.2	245
2	9.30	4.43	15.2	101.0	250
3	9.45	4.46	15.0	101.1	257
4	9.65	4.42	15.1	100.8	245
5	9.55	4.46	15.1	100.8	243
6	9.48	4.38	15.1	100.8	254
Mean	9.47	4.42	15.1	100.9	249
CV (%)	1.276	1.009	0.5	0.17	2.3
Former Calibration Factor (%)	92.95	97.08	96.01	101.11	101.42
New Calibration Factor (%)	97.01	98.72	96.49	100.25	97.59

Date: 27-01-2023

Substance: SC-CAL Calibrator

Calibration Performed by

Name: Mahesh CN

Signature:

