



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JUNE 2023



PC-1024

Lab Name MEDRAY CLINICS PVT LTD Lab No 17475
 Constituent Group Chemistry I Date of Result Entered : 16/06/2023
 PT item Lyophilized human serum based Date of Report Published : 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE	GOD-POD	Any Analyser (Automation / Semi Automation)	1157	123.88	6.67	8.26	138 mg/dL	1.71	0.49
2	UREA	Urease UV / GLDH	Any Analyser (Automation / Semi Automation)	759	48.17	9.57	4.61	46.9 mg/dL	-0.28	0.33
3	CREATININE	Jaffes End point	Any Analyser (Automation / Semi Automation)	332	2.01	9.39	0.19	2.49 mg/dL	2.54	0.02
4	T.BILIRUBIN	Diazonium salt (Colorimetric) / Jendrassik	Any Analyser (Automation / Semi Automation)	1028	2.93	12.68	0.37	3.39 mg/dL	1.24	0.02
5	T-PROTEIN	Biuret - Colorimetric	Any Analyser (Automation / Semi Automation)	1151	5.92	7.83	0.46	5.48 g/dL	-0.95	0.03
6	ALBUMIN	BCG - colorimetric	Any Analyser (Automation / Semi Automation)	829	3.54	7.30	0.26	3.56 g/dL	0.08	0.02
7	CALCIUM	Arsenazo III	Any Analyser (Automation / Semi Automation)	968	10.07	7.04	0.71	10.26 mg/dL	0.27	0.05
8	URIC ACID	Enzymatic / Uricase Colorimetric	Any Analyser (Automation / Semi Automation)	1029	5.37	15.52	0.83	4.26 mg/dL	-1.33	0.05
9	CHOLESTEROL	CHOD-PAP	Any Analyser (Automation / Semi Automation)	1201	124.53	6.12	7.62	134.11 mg/dL	1.26	0.44
10	TRIGLYCERIDE	GPO-PAP / Enzymatic Colorimetric / End Point	Any Analyser (Automation / Semi Automation)	1112	176.08	8.42	14.82	168.5 mg/dL	-0.51	0.89
11	HDL	Direct method / Enzymatic colorimetric	Any Analyser (Automation / Semi Automation)	862	29.06	13.00	3.78	31.7 mg/dL	0.70	0.26
12	AST	UV kinetic(with & without PLP (P-5-P))	Any Analyser (Automation / Semi Automation)	1094	66.44	12.71	8.45	77 U/L	1.25	0.51
13	ALT	UV kinetic(with & without PLP (P-5-P))	Any Analyser (Automation / Semi Automation)	1006	42.17	19.32	8.15	54.1 U/L	1.46	0.51
14	ALP	PNP AMP kinetic	Any Analyser (Automation / Semi Automation)	938	105.30	10.80	11.37	189 U/L	7.36	0.74

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
 MEDRAY CLINICS PVT LTD
 NO.962, 12TH MAIN ROAD, NEAR RELIANCE DIGITAL HALL 2ND STAGE, INDIRA NAGAR
 BANGALORE
 KARNATAKA560008

Investigation checklist/Form

Survey information: CMC Vellore June-2023

Survey Name	Chemistry	Analyzer name/Model	Autochem Expert
Date survey received	04/07/2023	Date of analysis performed	16/06/2023
Date survey result submitted	16/06/2023	Date of report receipt	04/07/2023
Investigation performed by	Prashanth	Date of evaluation	07/07/2023
Unacceptable parameter Name:		Date of retesting:	
Specimen	Analyte	Reported value	Repeated value
ATP <u>ATP</u> June 2023	ATP	189 U/L	-
			Intended/peer group value
			105.30 U/L

Root cause Analysis

Clerical	Yes	No	NA
1. Was the results correctly transcribed from instrument readout or report?	<input checked="" type="checkbox"/>		
2. Was the correct instrument /method reagent reported on the result form?	<input checked="" type="checkbox"/>		
3. Does the result reported on the result form match the result found on the proficiency testing evaluation report ?	<input checked="" type="checkbox"/>		

Procedural

1. Was the written procedure followed?	<input checked="" type="checkbox"/>		
2. Were the reagents within their open stability limit during analysis?	<input checked="" type="checkbox"/>		
3. Were Quality Control results acceptable and without bias?	<input checked="" type="checkbox"/>		
4. Were dilutions performed correctly?			<input checked="" type="checkbox"/>

Analytical

1. Was the most recent calibration acceptable and within established limits at the time of testing?	<input checked="" type="checkbox"/>		
2. Does a review of the past proficiency testing results indicate evenly distributed data without bias?		<input checked="" type="checkbox"/>	
3. Was the intended result within measuring range for the instrument?	<input checked="" type="checkbox"/>		
4. Was instrument maintenance performed on schedule?	<input checked="" type="checkbox"/>		
5. Does a review of records indicate that there were no related instrument test problems noted prior to or after the proficiency testing as performed?	<input checked="" type="checkbox"/>		

PT /EQAS material

1. Was proficiency testing material received in the laboratory within an appropriate time after shipment?	<input checked="" type="checkbox"/>		
2. Was proficiency testing material received at the appropriate temperature?	<input checked="" type="checkbox"/>		
3. Were results graded in the appropriate peer group based on the method reported on the result form?	<input checked="" type="checkbox"/>		

Conclusion /Summary: Random error suspected, to observe for

Type of error next survey result.

Method related	Survey evaluation problem	
Technical process related	<input checked="" type="checkbox"/>	Other (define below)
Clerical		

Preventive actions (If any) - to perform re calibration with fresh calibrators.

Review and approval:



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JUNE 2023



PC-1024

Lab Name	MEDRAY CLINICS PVT LTD	Lab No	17475
Constituent Group	HbA1c	Date of Result Entered :	16/06/2023
PT item	Lyophilized human whole blood based	Date of Report Published :	04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	HbA1c	HPLC-Ion Exchange	Any Analyser	164	6.34	8.79	0.56	6.1 %	-0.43	0.09

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

MEDRAY CLINICS PVT LTD
NO.962, 12TH MAIN ROAD, NEAR RELIANCE DIGITAL HALL 2ND STAGE, INDIRA NAGAR
BANGALORE
KARNATAKA560008

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****