



PROFICIENCY TESTING REPORT
ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME
 NABL accredited program as per ISO/IEC 17043:2010 standard
 Organized By Department of Hematology, AIIMS, New Delhi-110029



Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens

EQAP CODE No. : 4437

Distribution No.: 159-L

Month/Year: April/2023

Instrument ID: 20207

Name & Contact No. of PT Co-ordinator: Dr. Seema Tyagi (Prof.), Hematology, AIIMS, Delhi,
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue & status of the report: 13-06-2023[Final].

CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 ³ /µl	1	4.15	4.06	8.21	7.96	0.031	0.31	0.09	0.1	0.005	-0.11
RBC x10 ⁶ /µl	1	4.33	4.33	8.66	8.96	0.010	-1.09	0	0.05	0.003	-1.12
Hb g/dl	1	13.2	13.2	26.4	26.3	0.028	0.12	0	0.1	0.007	-1.15
HCT%	1	44.6	44.3	88.9	85.75	0.229	0.48	0.3	0.5	0.025	-0.45
MCV-fl	1	103	102.3	205.3	190.85	0.466	0.98	0.7	0.2	0.020	1.35
MCH-Pg	1	30.5	30.5	61	58.7	0.071	1.07	0	0.2	0.011	-0.90
MCHC-g/dl	1	29.8	29.6	59.4	60.8	0.162	-0.30	0.2	0.3	0.016	-0.34
Plt. x10 ³ /µl	1	130	124	254	251.5	1.160	0.08	6	5	0.327	0.17
Retic %	2										

P.S . Assesment

YOUR REPORT			CONSENSUS REPORT
DLC%	3	Nrbcs= , Poly= L=, E=, Mono/Promono= , B1= P.M.=, Mye=, Meta=, Other=	Lymp: 80-89, Poly: 9-15, Mono: 1-2, nRBC/blast/Eosino/Myelo/Meta: 0-1
RBC Morphology	3		Predominantly: Normocytic/Normochromic; Moderate: Anisocytosis, Microcytosis, Hypochromia; Mild: Macrocytosis, Poikilocytosis
Diagnosis	3		Chronic Lymphocytic Leukemia (CLL)

COMBINED DATA VALUES OF TOTAL PARTICIPANTS

Test parameters	S.No.	Total participants covered in the current dist. 159--L	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
WBC x10³/µl	1	346	346	84.68	90.17	4.05	3.47	11.27	6.36
RBC x10⁶/µl	1	346	346	87.57	91.33	8.09	3.76	4.34	4.91
Hb g/dl	1	346	346	90.17	89.31	5.78	6.65	4.05	4.04
HCT%	1	346	346	91.91	91.04	4.91	6.65	3.18	2.31
MCV-fl	1	346	346	93.64	91.33	3.18	4.34	3.18	4.33
MCH-Pg	1	346	346	93.06	92.2	4.62	2.89	2.32	4.91
MCHC-g/dl	1	346	346	92.77	90.17	5.49	3.47	1.74	6.36
Plt. x10³/µl	1	346	346	92.49	90.17	4.91	4.05	2.6	5.78
ReticCount%	2	346	223	92.38	91.48	7.17	2.69	0.45	5.83
PS Assessment	3	346	212	Satisfactory :95.96%, Borderline Sat. :3.18%, Unsatisfactory :0.86%					

***Comments:**

1). Among Lab (EQA) : PS Diagnosis not reported, remaining results acceptable

2). Within Lab (IQA) : Precision acceptable.

Note-1: EQA (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

IQA (Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

Note-2: Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

Note-3: Z score 0 to ±2: Acceptable, Z score ±2 to ±3 :Warning Signal, Z score > ±3 : Unacceptable [As per ISO/IEC 13528:2015 standard]

Note-4: Z score value between“0 to ±2” are texted in green colour. Z score value between“±2 to ±3” are texted in orange colour. Z score value > ±3 are texted in red colour.

Note-5: Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ($\bar{x}-\bar{y}$) should be smaller than the check value (0.3*SDPA).

Note-6: ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

Note-7: Participants are free to use methods/analyzer of their own choice.

Note-8: Proficiency testing (PT) samples are sent quarterly to each participant.

Note-9: All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme’s website www.ishtmaiimseqap.com.

Note 10: Reports are kept confidential.

Report authorized by,



Dr. Seema Tyagi (Prof.)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----

**District Govt. Hospital Sehore Central Processing Lab Sewan River
Square Sehore MP 466001**

Check list of Investigation/Corrective Action taken for EQAS failure

Parameter: Retic% and PS Diagnosis (Slide)		Month: April 2023	
SN	Pre-analytical phase of analysis	Status	
1	Was the EQAS sample stored at the proper temperature following receipt	Yes	<input checked="" type="checkbox"/>
2	Was the test reagent stored correctly in appropriate temperature	Yes	<input type="checkbox"/>
3	Was both slides unbroken while received	Yes	<input type="checkbox"/>
Analytical phase of analysis			
1	Was the sample at room temperature	Yes	<input type="checkbox"/>
2	Was the person running the EQAS sample was trained	Yes	<input type="checkbox"/>
3	Was daily maintenance performed on the day that the EQAS sample was run	Yes	<input type="checkbox"/>
4	Was IQC within an acceptable range on the day that the EQAS sample was run	Yes	<input type="checkbox"/>
Post-analytical phase of analysis			
1	Was the result uploaded before the last date of submission	Yes	<input type="checkbox"/>
2	Have the results been reported correctly (Match instrument raw data)	Yes	<input type="checkbox"/>
3	Was the configuration correct (instrument, method and reagent)	Yes	<input type="checkbox"/>
4	Was microscopy done again (in case of unacceptable results in Microscopic examination)	No	<input type="checkbox"/>

Note: Retic% and PS Diagnosis not reported in parameter due to the microscopy late examine, result has been assesst by two slides by the pathologist, after root cause analysis has been done as per above checklist. The first result of retic% is 8.1 and the second result is 7.4 and sum of two result 15.5.

DLC%: Nrbs=01, Poly=10.0 L=83, E=01, Mono/Promo=03, B1=02, P.M.=00, Mye=00, Meta=00, Other=00.
RBC Morphology: Normocytic Normochromic Diagnosis: Chronic Lymphocytic Leukemia CLL, now the Retic% and PS Diagnosis are acceptable.


डॉ. आनन्द बर्म
 पैथोलॉजी विशेषज्ञ
 जिला बिकित्सालय सीहोर

Reviewed by:

Date: 31-7-23

Dear Sir/Mo'am,

Lab can not submit subsequent PT report because the AHMS EQAS specimens received quarterly. for this reason Lab will received next subsequent PT report in September month.

SAMPLE NO.1: PERIPHERAL BLOOD

NO.1 WBC(Higher Value: On the scale of ** x10³/μl)

8.17

NO.2 WBC(Lower Value:On the scale of ** x10³/μl)

8.04

NO.1 RBC (Higher Value:On the scale of ** x10⁶/μl)

4.90

No.2 RBC(Lower Value:On the scale of ** x10⁶/μl)

4.89

No.1 Hb(Higher Value:On the scale of ** g/dl)

14.0

No.2 Hb.(Lower Value:On the scale of ** g/dl)

14.0

No.1 Hct(Higher Value:On the scale of ** %)

44.9

No.2 Hct(Lower Value:On the scale of ** %)

44.6

No.1 MCV(Higher Value:On the scale of ** fl)

91.6

No.2 MCV(Lower Value:On the scale of ** fl)

91.2

No.1 MCH(Higher Value:On the scale of ** pg)

28.6

No.2 MCH(Lower Value:On the scale of ** pg)

28.6

No.1 MCHC(Higher Value:On the scale of ** g/dl)

31.4

No.2 MCHC(Lower Value:On the scale of ** g/dl)

31.2

No.1 Plt(Higher Value:On the scale of ** x10³/μl)

153

No.2 plt(Lower Value:On the scale of ** x10³/μl)

136

SAMPLE NO.2: RETICULOCYTE

No.1 retic(Higher Value:On the scale of ** %)

5.5

No.2 retic(Lower Value:On the scale of ** %)

5.0

SAMPLE NO.3: PERIPHERAL SMEAR

nRBC

00

Poly/band/stab

00

Lympho/prolympho

20

Eosino

02

BASO

00

Mono/Promono

02

Blast ✓

60

Promyelo

02

Myelo

00

Meta

01

Others

HAND MIRROR OR SMUDGE CELL

RBC morphology

MICROCYTIC HYPOCHROMIC

Diagnosis

ACUTE LEUKEMIA

Date of receipt of specimens *

2023-07-25

Date of processing of specimens *

2023-08-01

Quality of Specimens

No.1

GOOD

No.2

GOOD

No.3

GOOD

Suggestion/Feedback

Suggestion/ Feedback

SUBMIT