


# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



## Calibration Certificate

<b>CERTIFICATE NO.</b>	: 230626A/N-111	<b>Page</b>	: 1 of 1
<b>CALIBRATED ON.</b>	: 28 / 06 / 2023	<b>Ref. Std.</b>	: IS:12508:1988
<b>CALIBRATION DUE.</b>	: 27 / 06 / 2024	<b>C.P. No.</b>	: ACS/CP/10
<b>SRF No. &amp; DATE</b>	: 230626A/N Dt. 26.06.2023	<b>Phy. Inspection</b>	: O.K.
<b>ULR No. &amp; ISSUE DATE</b>	: CC342723000000627F Dt. 30.06.2023	<b>Cal. At</b>	: Site / Lab
<b>CALIBRATED FOR</b>	: M/s Redcliffe Lifetech Pvt. Ltd. A-9 Civil Lines, GT Road, Opp Civil Hospital, Bhatinda, Punjab -151001		
<b>ITEM CALIBRATED</b>	: Centrifuge	<b>Make</b>	: Remi
		<b>Range</b>	: (0 to 6000) RPM
		<b>L. Count</b>	:
		<b>ID No.</b>	: RCL/BIO/Cent-01
		<b>Location</b>	: Clinical Biochemistry
		<b>S. No.</b>	: ZIEN-24034
<b>ENVIRONMENT CONDITION:</b>	Temp. { 25°C ± 3°C } R.H. {40% to 60%}		
<b>STANDARD INSTRUMENTS USED :</b>			
<b>INSTT. NAME</b>	<b>ID No. / S. No.</b>	<b>TRACEABILITY</b>	<b>CERTIFICATE No.</b>
<b>DUE DATE</b>			
Tachometer	ACS/DTM/01/ 211-0155V	Godrej	C-220912-6-1
			12/09/2023
<b>RESULTS</b>	: Mechanical Discipline		
* U. C. INSTT.	Standard Measured	Error	
READS (RPM)	(RPM)	(RPM)	
1000	1001.42	- 1.42	
2000	2002.36	- 2.36	
4000	4005.32	- 5.32	
6000	6006.38	- 6.38	
(average of readings)			

**REMARKS** : The Expanded uncertainty of measurement at 95% confidence level is ± 3.60 rpm of reading at coverage Factor k=1.96.

\*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

