

ADVANCE CALIBRATION SERVICES





(ISO: 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob.: +91-9891121066, 9868309579 • E-mail: advance.cal@gmail.com

Calibration Certificate

CERTIFICATE NO.	: 2306264			Page	: 1 of 1			
CALIBRATED ON.	2023		Ref. Std.	: ISO:8655:6				
CALIBRATION DU	2024			C.P. No.	No. : ACS/CP/08			
SRF No. & DATE	/N Dt. 26.06.2023 .			Phy. Inspection : O.K.				
ULR No. & ISSUE DATE : CC3427		23000000617F Dt. 30.06.2023			Cal. At	: Site / Lab		
CALIBRATED FOR	A-9	cliffe Lifetech Civil Lines, C itinda, Punjab	ST Road, Op	p Civil I	Hospital,			
ITEM CALIBRATED	: Micro Pi	pette						
	Make	: Thermo Scientific			Range : (2 to 20) μl ID No. : RCL/HEM/PIP-01			
	t :	:		-01				
Location : Haematology				S. No. : TW08221				
ENVIRONMENT CO	ONDITION: Te	emp. { 25°C <u>+</u>	3°C} R.H	I. {40%	to 60%}			
STANDARD INSTR	UMENTS USED	:	***************************************					
INSTT. NAME	ID No. / S. No.		TRACEABILITY		CERTIFICATE I	No. DUE	DATE	
Weighing Balance	ACS/DB/02/ D449927064		BlueBox		BBT/398/JAN/2	05/01	/2024	
RESULTS :	Mechanical	Discipline						
Nominal Capacity	Graduation	Nominal \	'alue Vol		ume at 20 °C	Error	Error	
(µI)	(µI)	(µl)			(µI)	(µI)	(µI)	
		10			10.08	(-) 0.08	(-) 0.08	
20								
		20	The state of the s		20.12		(-) 0.12	

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 0.1 µl at coverage factor k=1.96.

Under Calibration.

Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.

The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above. This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.

The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.

Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use

Confirmity statement not provided since not required.



