



ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 230626A/N-105	Page	: 1 of 1	
CALIBRATED ON.	: 26 / 06 / 2023	Ref. Std.	: ISO:8655:6	
CALIBRATION DUE.	: 25 / 06 / 2024	C.P. No.	: ACS/CP/08	
SRF No. & DATE	: 230626A/N Dt. 26.06.2023	Phy. Inspection	: O.K.	
ULR No. & ISSUE DATE	: CC342723000000621F Dt. 30.06.2023	Cal. At	: Site / Lab	
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. A-9 Civil Lines, GT Road, Opp Civil Hospital, Bhatinda, Punjab -151001			
ITEM CALIBRATED	: Micro Pipette			
	Make	: Thermo Scientific	Range	: (10 to 100) µl
	L. Count	:	ID No.	: RCL/BIO/PIP-05
	Location	: Clinical Biochemistry	S. No.	: TW06256
ENVIRONMENT CONDITION:	Temp. { 25°C ± 3°C } R.H. {40% to 60%}			
STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/ D449927064	BlueBox	BBT/398/JAN/23	05/01/2024
RESULTS :				Mechanical Discipline
Nominal Capacity (µl)	Graduation (µl)	Nominal Value (µl)	Volume at 20 °C (µl)	Error (µl)
		10	10.10	(-) 0.10
100		50	50.20	(-) 0.20
		100	100.42	(-) 0.42

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 0.1 µl at coverage factor k=1.96.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmary statement not provided since not required.

Prepared by

....End of Calibration Certificate....

