


# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



## Calibration Certificate

<b>CERTIFICATE NO.</b>	: 230626A/N-106	<b>Page</b>	: 1 of 1
<b>CALIBRATED ON.</b>	: 26 / 06 / 2023	<b>Ref. Std.</b>	: ISO:8655:6
<b>CALIBRATION DUE.</b>	: 25 / 06 / 2024	<b>C.P. No.</b>	: ACS/CP/08
<b>SRF No. &amp; DATE</b>	: 230626A/N Dt. 26.06.2023	<b>Phy. Inspection</b>	: O.K.
<b>ULR No. &amp; ISSUE DATE</b>	: CC342723000000622F Dt. 30.06.2023	<b>Cal. At</b>	: Site / Lab
<b>CALIBRATED FOR</b>	: M/s Redcliffe Lifetech Pvt. Ltd. A-9 Civil Lines, GT Road, Opp Civil Hospital, Bhatinda, Punjab -151001		
<b>ITEM CALIBRATED</b>	: Micro Pipette		
	<b>Make</b> : Thermo Scientific	<b>Range</b>	: (100 to 1000) µl
	<b>L. Count</b> :	<b>ID No.</b>	: RCL/BIO/PIP-06
	<b>Location</b> : Clinical Biochemistry	<b>S. No.</b>	: TW06215
<b>ENVIRONMENT CONDITION:</b>	Temp. { 25°C ± 3°C } R.H. {40% to 60%}		
<b>STANDARD INSTRUMENTS USED :</b>			
<b>INSTT. NAME</b>	<b>ID No. / S. No.</b>	<b>TRACEABILITY</b>	<b>CERTIFICATE No.</b>
Weighing Balance	ACS/DB/02/ D449927064	BlueBox	BBT/398/JAN/23
			<b>DUE DATE</b>
			05/01/2024
<b>RESULTS :</b>			
<b>Mechanical Discipline</b>			
<b>Nominal Capacity (µl)</b>	<b>Graduation (µl)</b>	<b>Nominal Value (µl)</b>	<b>Volume at 20 °C (µl)</b>
			<b>Error (µl)</b>
		100	100.36
1000		500	501.66
		1000	1003.30
			(-) 0.36
			(-) 1.66
			(-) 3.30

**REMARKS** : The Expanded uncertainty of measurement at 95% confidence level is ± 0.6 µl at coverage factor k=1.96.

\*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

