



	A	B	C	Σ	
1	HUMALYZER -3000 CALIBRATION RAW DATA				
2	PARAMETERS	STD CONCENTRAION	FACTOR OBTAINED	FACTOR RANGE	REMARK
3	GLUCOSE	100 mg/dl	269.5	220-300	VALID
4	CHOLESTEROL	200 mg/dl	524.9	450-650	VALID
5	TRIGLYCERIDES	200 mg/dl	689.696	550-750	VALID
6	CREATININE	2 mg/dl	36.29	25-45	VALID
7	CALCIUM	10 mg/dl	17.12	15-20	VALID
8					
9					
10					

CERTIFICATE OF CALIBRATION

To,
City Lab,
Thalassery

This is to certify that the instrument Humalyzer 3000 bearing Serial No. 3500-5449 manufactured by Human GMBH of Germany is under Calibration. Quality control check up of the instrument has been the responsibility of MEDSOURCE OZONE BIOMEDICALS PVT. LTD.

Date of Calibration : 13th April, 2023

Calibration & quality control checks as recommended by the manufacturers in operator manual are being performed regularly and satisfactorily.

This calibration certificate is valid up to 12th April 2024

For Medsource Ozone Biomedicals Pvt. Ltd.



Reshma UR
Service Coordinator
+91-9650064495

Medsource Ozone Biomedicals Pvt. Ltd.

Head Office: Plot No. 109, HSIIDC, Sector-31, Faridabad-121003, (Haryana), INDIA
Registered Office: 238, Sant Nagar, East of Kailash, New Delhi-110065, INDIA
Telephone: +91-129-4286600 | E-mail: mail@ozonebio.com | Website: www.ozonebio.com