



**PROFICIENCY TESTING REPORT**  
 ISHM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME  
 NABL accredited program as per ISO/IEC 17043:2010 standard  
 Organized By Department of Hematology, AIIMS, New Delhi-110029



*Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens*

EQAP CODE No. : 2180

Distribution No.: 158-F

Month/Year: December/2022

Instrument ID: A6927

Name &amp; Contact No. of PT Co-ordinator: Dr. Seema Tyagi (Prof.), Hematology, AIIMS, Delhi,

Tel: 9013085730, E-Mail: accuracy2000@gmail.com

Date of issue &amp; status of the report: 23-01-2023[Final]

### CBC and Retic Assessment

| Test Parameters           | S.No. | Among Lab (Accuracy Testing) |               |                             |   |                                |         | Within Lab (Precision Testing)  |   |                                |         |
|---------------------------|-------|------------------------------|---------------|-----------------------------|---|--------------------------------|---------|---------------------------------|---|--------------------------------|---------|
|                           |       | Your Result 1                | Your Result 2 | Your Results Sum of 2 Value | Consensus result sum of 2 values (Assigned Value) | Uncertainty of Assigned Values | Z Score | Yours Results Diff. of 2 Values | Consensus Result Diff. of 2 values (Assigned Value) | Uncertainty of Assigned Values | Z Score |
| WBC x10 <sup>3</sup> /µl  | 1     | 10.2                         | 9.4           | 19.6                        | 18.2  | 0.0730                         | 0.67    | 0.8                             | 0.12  | 0.0090                         | 1.34    |
| RBC x10 <sup>6</sup> /µl  | 1     | 4.53                         | 4.5           | 9.03                        | 9.11  | 0.0100                         | -0.33   | 0.03                            | 0.04  | 0.0030                         | -0.27   |
| Hb g/dl                   | 1     | 12.7                         | 12.5          | 25.2                        | 25.8  | 0.0220                         | -1.16   | 0.2                             | 0.1   | 0.0080                         | 1.35    |
| HCT%                      | 1     | 41.3                         | 41.1          | 82.4                        | 84.35   | 0.2260                         | -0.33   | 0.2                             | 0.4   | 0.0260                         | -0.54   |
| MCV-f                     | 1     | 91.3                         | 91.2          | 182.5                       | 185.4   | 0.4410                         | -0.23   | 0.1                             | 0.2   | 0.0210                         | -0.34   |
| MCH-Pg                    | 1     | 28                           | 27.8          | 55.8                        | 56.4  | 0.0550                         | -0.45   | 0.2                             | 0.2   | 0.0140                         | 0.00    |
| MCHC-g/dl                 | 1     | 30.8                         | 30.4          | 61.2                        | 60.7  | 0.1560                         | 0.11    | 0.4                             | 0.2   | 0.0160                         | 0.90    |
| PLT. x10 <sup>3</sup> /µl | 1     | 198                          | 196           | 394                         | 373   | 1.71                           | 0.46    | 2                               | 5   | 0.32                           | -0.58   |
| Retic %                   | 2     | 12                           | 11            | 23                          | 27  | 0.63                           | -0.23   | 1                               | 1   | 0.06                           | 0.00    |

### P.S. Assessment

| YOUR REPORT    |   | CONSENSUS REPORT  |
|----------------|---|---|
| DLC%           | 3 | Nrbc=05, Poly=44, L=07, E=02, Mono/Promono=01, B1=06 P.M.=02, Mye=27, Meta=10, Other=   |
| RBC Morphology | 3 | Predominantly: Normocytic/Normochromic, Moderate: Anisocytosis, hypochromia, Microcytosis; Mild: Macrocytosis, Poikilocytosis |
| Diagnosis      | 3 | Chronic Myeloid Leukemia (Chronic Phase)  |

CAPA Attached

*[Signature]*

## COMBINED DATA VALUES OF TOTAL PARTICIPANTS

| Test parameters           | S.No. | Total participants covered in the current dist. 158--F | Total No. responded | % of Labs with Z Score 0-2  |            | % of Labs with Z Score 2-3 |            | % of Labs with Z Score >3 |            |
|---------------------------|-------|--|---------------------|---|------------|----------------------------|------------|---------------------------|------------|
|                           |       |  |                     | Among labs  | Within lab | Among labs                 | Within lab | Among labs                | Within lab |
| WBC x10 <sup>3</sup> /µl  | 1     | 296  | 295                 | 87.46   | 88.81      | 5.42                       | 5.08       | 7.12                      | 6.11       |
| RBC x10 <sup>6</sup> /µl  | 1     | 296  | 296                 | 83.78   | 92.23      | 9.46                       | 3.38       | 6.76                      | 4.39       |
| Hb g/dl                   | 1     | 296  | 296                 | 82.77   | 89.53      | 8.45                       | 6.08       | 8.78                      | 4.39       |
| HCT%                      | 1     | 296  | 294                 | 95.92   | 94.22      | 3.06                       | 2.04       | 1.02                      | 3.74       |
| MCV-fl                    | 1     | 296  | 294                 | 97.62   | 93.2       | 1.7                        | 2.04       | 0.68                      | 4.76       |
| MCH-Pg                    | 1     | 296  | 294                 | 85.03   | 91.16      | 7.14                       | 3.4        | 7.83                      | 5.44       |
| MCHC-g/dl                 | 1     | 296  | 294                 | 97.28   | 92.86      | 1.36                       | 3.74       | 1.36                      | 3.4        |
| Plt. x10 <sup>3</sup> /µl | 1     | 296  | 294                 | 94.22   | 91.84      | 4.76                       | 4.42       | 1.02                      | 3.74       |
| ReticCount%               | 2     | 296  | 265                 | 92.83   | 93.58      | 4.53                       | 2.64       | 2.64                      | 3.78       |
| PS Assessment             | 3     | 296  | 265                 | Satisfactory :92.21%, Borderline Sat. :3.05%, Unsatisfactory :4.74% |            |                            |            |                           |            |

## Comments:

1). Among Lab (EQA) : Results acceptable.

2). Within Lab (IQA) : Difference in the CBC measurement values for WBC unacceptable, may be due to random/human error.

**Note-1:** EQA (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

**IQA** ( Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

**Note-2:** Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

**Note-3:** Z score 0 to ±2: Acceptable, Z score ±2 to ±3 :Warning Signal, Z score > ±3 : Unacceptable [As per ISO/IEC 13528:2015 standard]

**Note-4:** Z score value between "0 to ±2" are texted in green colour. Z score value between "±2 to ±3" are texted in orange colour. Z score value > ±3 are texted in red colour.

**Note-5:** Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3\*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ( $\bar{x}-\bar{y}$ ) should be smaller than the check value (0.3\*SDPA).

**Note-6:** ISITM-AIIMS-EQAP does not subcontract any task of its scheme


**Note-7:** Participants are free to use methods/analyzer of their own choice.

**Note-8:** Proficiency testing (PT) samples are sent quarterly to each participant.

**Note-9:** All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website [www.ishtmaiimseqap.com](http://www.ishtmaiimseqap.com).

**Note 10:** Reports are kept confidential.

Report authorized by,



Dr. Seema Tyagi (Prof.)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----

# GLOBAL PATHOLOGY CENTRE, HARI NAGAR

## EQAS EVALUATION FORMAT

Doc. No. GPC/FR/65

|                          |                        |                            |    |                    |
|--------------------------|------------------------|----------------------------|----|--------------------|
| EQAS Sample details      | Bio-Rad/CMC/AIIMS EQAS | Sample: Lyophilized/Liquid |    |                    |
| EQAS Lot No.             | 158 F                  | Received on:               |    | Received By: Babli |
| CAPA Date:               | 14/4/2023              | Cycle No:                  |    | Month: December    |
| Date reconstituted:      | NA                     | Reconstituted By:          | NA | 2022               |
| Date & Time of Analysis: |                        |                            |    |                    |
| Aliquot preserved on :   |                        |                            |    |                    |
| Results received on :    | 25/01/2023             |                            |    |                    |

**Criteria for outlier SDI/Z score : > 3.0 and Repetition of score >2.0(2 Consecutive cycle)**

|                    |     |              |        |         |
|--------------------|-----|--------------|--------|---------|
| Parameter outside: | TLC | SDI/Z score: | 8.34   | Method: |
| Peer group result: |     | Lab Result:  |        |         |
| Internal QC data:  | Low | High         | Normal |         |
| Date:              |     |              |        | ✓       |
| U chart findings:  | OK  |              |        |         |

**Cause of error detected in Lab (Pls Tick)**

- 1) **Method problem** : Calibration failure, Inadequate maintenance, deterioration of reagent, OK  
 Deterioration of component, environmental factors. OK
- 2) **Procedure problem**: Incorrect dilution, Incorrect calculation. OK
- 3) **PT sample handling**: Cold chain & storage, Reconstitution, Mixing error. OK
- 4) **Incorrect Reporting**: Incorrect conversion of units, Incorrect method, transcription errors. OK
- 5) **Problem with PT material**: Unstable analyte, Interfering substance. OK

|                  |  |                     |  |              |   |
|------------------|--|---------------------|--|--------------|---|
| Systematic Error |  | Transcription Error |  | Random error | ✓ |
|------------------|--|---------------------|--|--------------|---|

**Corrective Action Taken:**

|                       |  |             |               |              |
|-----------------------|--|-------------|---------------|--------------|
| Retained sample run : |  | IQC:        | Done for P/C. | PRECISION: L |
| New Z score:          |  | %Variation: | Satisfactory  | 14/4/2023    |

**Preventive action taken:**

will take care during sample processing in future. Satisfactory

*[Handwritten signature]*

*[Handwritten signature]*  
14/4/2023

GLOBAL PATHOLOGY CENTRE, Hari Nagar

PRECISION CHECK

HAEMATOLOGY-H560

| SL No. | DATE        | TLC         | RBC          | HB           | PCV          | MCV           | MCH          | MCHC        | PLT          |
|--------|-------------|-------------|--------------|--------------|--------------|---------------|--------------|-------------|--------------|
| 1      | 14/04/2023  | 5.3         | 3.47         | 11.6         | 35.2         | 101.4         | 33.4         | 33          | 146          |
| 2      |             | 5.6         | 3.43         | 11.8         | 35           | 102           | 34.4         | 33.7        | 140          |
| 3      |             | 5.2         | 3.44         | 11.7         | 34.9         | 101.5         | 34           | 33.5        | 139          |
| 4      |             | 5.4         | 3.45         | 11.8         | 35.1         | 101.5         | 34.2         | 33.6        | 134          |
| 5      |             | 5.4         | 3.44         | 11.8         | 35           | 101.7         | 34.3         | 33.7        | 139          |
|        | <b>MEAN</b> | <b>5.38</b> | <b>3.446</b> | <b>11.74</b> | <b>35.04</b> | <b>101.62</b> | <b>34.06</b> | <b>33.5</b> | <b>139.6</b> |
|        | <b>SD</b>   | 0.15        | 0.02         | 0.09         | 0.11         | 0.24          | 0.40         | 0.29        | 4.28         |
|        | <b>CV%</b>  | 2.8         | 0.4          | 0.8          | 0.3          | 0.2           | 1.2          | 0.9         | 3.1          |

Acceptance Criteria: <10% of CV

Prepared By: Quality Manager

Approved By: Lab Director

