



www.dlabsci.com
YE229AX0539447 CE




QUALITY CONTROL CERTIFICATE
QUALITÄTSKONTROLLZERTIFIKAT
CERTIFICAT DE CONTRÔLE QUALITÉ
CERTIFICADO DE CONTROL DE CALIDAD
CERTIFICATO DI CONTROLLO QUALITÀ

Pipette 100-1000ul

Serial Number/Serien-Nr/N° de série/Número de serie/Numero di Soric: YE229AX0539447

Mode/Programm/Modo/Funzione: P Volume/Volumen: 200.00 ul
Mean/Mittelwert/Moyenne/Significado/Medio: 199.08 ul
Inaccuracy/Unrichtigkeit/Erreur de justesse/Inexatitud/Inaccuratezza: -0.46 %
Imprecision/Unpräzision/Erreur de répétabilité/Imprecisión/Imprecisione: 0.18 %

31/3/2023



Date/Datum/Date/Fecha/Data: 2022-09-13 14:03:52
According to/Nach/Selon/Segun/In base a: ISO8655 DIN 12650, 22°C

Tested by/Prüfer/Testée par/Comprobado por/Testado da: Fern

CE ISO 9001/13485
CERTIFIED



QUALITY CONTROL CERTIFICATE
QUALITÄTSKONTROLLZERTIFIKAT
CERTIFICAT DE CONTRÔLE QUALITÉ
CERTIFICADO DE CONTROL DE CALIDAD
CERTIFICATO DI CONTROLLO QUALITÀ

Pipette 10-100ul

Serial Number/Serien-Nr/N° de série/Número de serie/Numero di Soric: YE221AW0155238

Mode/Programm/Modo/Funzione: P Volume/Volumen: 25.00 ul
Mean/Mittelwert/Moyenne/Significado/Medio: 25.23 ul
Inaccuracy/Unrichtigkeit/Erreur de justesse/Inexatitud/Inaccuratezza: 0.92 %
Imprecision/Unpräzision/Erreur de répétabilité/Imprecisión/Imprecisione: 0.45 %

31/3/2023



Date/Datum/Date/Fecha/Data: 2022-01-17 10:06:27
According to/Nach/Selon/Según/In base a: ISO8655 DIN 12650; 20°C
Tested by/Prüfer/Testée par/Comprobado por/Testado da: Fern

CE ISO 9001/13485
CERTIFIED