

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JUNE 2023



PC-1024

Lab Name J.P HOSPITAL , BHOPAL Lab No 7791
Constituent Group HbA1c Date of Result Entered : 15/06/2023
PT item Lyophilized human whole blood based Date of Report Published : 04/07/2023

| Sl.No | Analyte | Method / Principle Name | Analyzer Name | No of Participants | DV | Participants | | Your Value | SDI | U |
|-------|---------|-------------------------|---------------------------------------|--------------------|------|--------------|------|------------|-------|------|
| | | | | | | CV | SD | | | |
| 1 | HbA1c | HPLC-Ion Exchange | Biorad (D10 / D100 / Variant Turbo) | 715 | 6.80 | 2.65 | 0.18 | 6.7 % | -0.56 | 0.01 |

| SDI Range | Interpretation |
|---------------------------------|--|
| Within -1.00 to +1.00 | Excellent. |
| Within ± 1.01 to ± 2.00 | Good. |
| Within ± 2.01 to ± 2.99 | Accept with caution. Warning Signal. |
| Beyond ± 3.0 | Unacceptable performance. Action Signal. |

LAB ADDRESS :
J.P HOSPITAL , BHOPAL
TULSHI NAGAR, BHOPAL
BHOPAL
MADHYA PRADESH 462004

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

