

CAC

**BIO-RAD**

BIO-RAD LABORATORIES (INDIA) PRIVATE LIMITED

Site name:	District hospital Bhopal
Address:	1250 Link Road No. 2 Tulsi nagar
Province-Postal code	462001
Contact Name	Mr. Abhinav

**Calibration Certificate**

Date	01-09-2022 <small>(dd mmmm, yyyy)</small>	Instrument Model	D-10 Hb Analyzer	Instrument Serial Number	DM0G519494
------	--	------------------	------------------	--------------------------	------------

We certify that the Bio-Rad instrument specified above and installed in your lab is calibrated and the performance was found satisfactory.

**High Pressure Pump Flow Rate Reference: Service Manual: Section 7.7.3**

Measured with <small>(Device manufacturer and model)</small>	Device Serial Number	Device Calibration date <small>(dd mmmm, yyyy)</small>
Glassco/measuring cylinder	11.20/2121	21/11/2020
Specified Range @ 1.5ml/Minute	Measured Value	
1.425 – 1.575 ml/minute	1.482	

**Column Holder temperature Reference: Service Manual: Section 7.9**

Measured with <small>(Device manufacturer and model)</small>	Device Serial Number	Device Calibration date <small>(dd mmmm, yyyy)</small>
Fluke - 179	28400220	23-01-22
Specified Range @ 35°C	Measured Value	
33.4°C – 36.4°C	34.9	

**Detector log value Reference: Service Manual: Section 7.2 or Field Bulletin D-10-081**

Measured with	Specified Range	Measured Value
Instrument Service Software	0.018 – 0.022 AU	0.0200

The next calibration is due on	01-09-2023 <small>(dd mmmm, yyyy)</small>
--------------------------------	--

For Bio-Rad Laboratories India Pvt Limited

Authorized Signatory:

*AB*  
01/09/22

**ABHINAV PUROHIT**  
(Division Lab Co-ordinator-BHOPAL)

POCT SCIENCE HOUSE PVT. LTD.  
Bhopal, India. Phone: 0755-8827251034

Calibration Certificate