



Quality Services & Laboratories

Plot No.10, Second Floor, D.S.I.D.C. Scheme - III
Okhla Industrial Area, Phase - II, New Delhi- 110020, India,
Tel. : +91 85953 43496
Email: corp@qsl.co.in ; Web.: www.qslglobal.com



CALIBRATION CERTIFICATE

Page 1 of 1

Certificate No. : QSL/DEL/CAL/2305315001.01
Date of Calibration: 31.05.2023
Date of Receipt of Instruments: 31.05.2023

SRF No. : QSL/2305315001
Suggested Due Date: 31.05.2024
Date of Issue: 01.06.2023

Customer Name & Address:

M/s: Jai Prakash Hospital Bhopal
(1st floor Central Processing Lab) Tulsī Nagar Bhopal-(M.P.)

Description of Instrument:

Name :	* Centrifuge Machine (RPM)	Range :	As Per Range
Make :	Laby	Least Count :	As Per Range
S. No. :	2022-23/500	Accuracy :	-----
Model No.	SP-8BL	Performed at :	Site
Location:	1 st floor Central Processing Lab	Visual Inspection :	OK
Customer ID:	JP/CPL/Centrifuge/12	Zero Error :	Nil

Environmental Condition:

Temperature :	25°C ± 5°C	Humidity :	45 % to 75%
---------------	------------	------------	-------------

Methodology of Calibration:

Work Instruction No.: QSL/WI/ET/23

Standard & Major Equipment (s) Used for calibration:

S. No.	Instrument Name	Make	S. No. /Model No.	Certificate No.	Calibrated By	Valid Upto
1.	Digital Techometer	-----	L142329	CC223923000000204F	Asian	15.02.2024

Calibration Result:

UUC Reading in RPM	Standard Reading in RPM	Error (RPM)	Uncertainty of Measurement:
1500	1500	00	± 1 %
2000	2001	-01	
3000	3001	-01	

(The reported expanded Uncertainty in measurement is stated as the standard uncertainty in measurement multiplied by the coverage factor k = 2, which for a normal distribution corresponds to a coverage probability of approximately 95%)

Note:

- *The calibration results reported in this calibration certificate are valid at the time of & under stated conditions of measurement.
- *This certificate cannot be reproduced except in full without our prior permission in writing.
- *This certificate refers only to the particular item (s) calibrated.
- *Laboratory standard are traceable to national standards.
- *UUC: Unit Under Calibration.
- *Is not in our NABL Scope Of Accreditation

Prepared By	Approved By
Customer Service Cell (Gajendra Singh)	CEO (Mridul Kohli)



QSL -1P-142