

Instrument SI No.	K6LC-13793
Calibration Date	09/08/2023
Suggested Due Date	08/08/2024
Sign.	



A-337, Sumel Business Park-4, Kalupur-Naroda Road, Amdupura-382345, Ahmedabad.
M. 9974712437, 9173769174 | E-mail : arshenterprise1992@gmail.com

Calibration Certificate

Discipline :	Mechanical	Certificate No.:	CC312623000019552F
--------------	------------	------------------	--------------------

1) Customer Information:

Customer Details:	M/s. Jay Prakash Hospital	SRF No.	AE/SRF/23/2323
	1st Floor Central Processing Lab	Receipt Date	09/08/2023
	Tulsi Nagar/Shiva Ji Nagar, Bhopal	Calibration Date	09/08/2023
	Madhya Pradesh 462023.	Due Date	08/08/2024
		Date of Issue	10/08/2023

2) Instrument Information:

Calibrated Item	Centrifuge Machine	Visual Inspection	Ok
Make/Model	Remi / ===	Type	===
Range	0 - 5000 Rpm	Least Count	1 RPM
Serial No./ID No.:	KGLC-13793 / JP/CPL/CENTRIFUGE/20	Calibration Performed	At Site
User Location	Lab	Parameter	Speed

3) Environmental Conditions:

Temperature	25 ± 5 °C	Calibration Ref Std No.	===
Relative Humidity	50 ± 20 % RH	Calibration Procedure No.	AE/CAL/M/SOP-27

4) Reference Standards Used:

Reference Standard	Serial No.	Certificate No.	Valid Up To	Traceability
Digital Tachometer	AE/M/28	PE23CC02645	22/05/2024	CC-3034

5) Calibration Results:

Reading Observed in UUC	Reading shown in Master Digital Tachometer	Error
1000 RPM	1000.8 RPM	0.8 RPM
2000 RPM	2001.4 RPM	1.4 RPM
3000 RPM	3003.5 RPM	3.5 RPM
4000 RPM	4004.6 RPM	4.6 RPM
5000 RPM	5005.7 RPM	5.7 RPM

6) Notes:

- For All Notes Kindly Refer Overleaf
- Expanded Uncertainty : ± 1.21 %

Calibrated By

Authorised Signatory

Yash Khatri
(Calibration Engineer)

Mansuri Md.Arif
(Quality Manager)



End of Certificate -

*