



Quality Services & Laboratories

Plot No.10, Second Floor, D.S.I.D.C. Scheme - III
Okhla Industrial Area, Phase - II, New Delhi- 110020, India,
Tel. : +91 85953 43496
Email: corp@qsl.co.in ; Web.: www.qslglobal.com



CC-2900

CALIBRATION CERTIFICATE

Unique Lab Report No. : ULR-CC290023000004057F
Certificate No. : QSL/DEL/CAL/23053101.03
Date of Calibration: 31.05.2023
Date of Receipt of Instruments: 31.05.2023

Page 1 of 1
SRF No. : QSL/23053101
Suggested Due Date: 31.05.2024
Date of Issue: 01.06.2023

Customer Name & Address:

M/s: Jai Prakash Hospital Bhopal
(1st floor Central Processing Lab) Tulsi Nagar Bhopal-(M.P.)

Description of Instrument:

Name :	Micro Pipette	Range :	5 – 50 µl
Make :	Dragon	Least Count :	0.1 µl
S. No. :	VE20AA50046293	Accuracy :	-----
Model No. :	-----	Performed at :	Lab
Location:	1 st floor Central Processing Lab	Visual Inspection :	OK
Customer ID:	JP/CPL/MP/Dragon/18	Zero Error :	Nil

Environmental Condition:

Temperature :	25°C ± 3 °C	Humidity :	40% to 60%
---------------	-------------	------------	------------

Methodology of Calibration:

Reference Standard: IS: 4787	Work Instruction No. :	QSL/WI/M/21
------------------------------	------------------------	-------------

Standard & Major Equipment (s) Used for calibration:

S.No.	Instrument Name	Make	S. No./Model No.	Certificate No.	Calibrated By	Valid Upto
1.	Std.Weight Box	Weightronics	-----	WMCL/F/2021-06/1310	Weightronics	10.06.2023

Calibration Result:

UUC Reading in µl	Standard Reading in µl	Error in µl	Uncertainty of Measurement
10	9.62584	0.37416	± 0.9 µl
20	19.59721	0.40279	
50	49.48671	0.51329	

(The reported expanded Uncertainty in measurement is stated as the standard uncertainty in measurement multiplied by the coverage factor k = 2, which for a normal distribution corresponds to a coverage probability of approximately 95%)

Note:

- *The calibration results reported in this calibration certificate are valid at the time of & under stated conditions of measurement.
- *This certificate cannot be reproduced except in full without our prior permission in writing.
- *This certificate refers only to the particular item (s) calibrated.
- *Laboratory standard are traceable to national standards.
- *UUC: Unit Under Calibration

Prepared By

Customer Service Cell (Gajendra Singh)

Approved By

CEO (Mridul Kohli)



QSL -1P-140

All business are undertaken subject to company standard general business conditions, which are available on request.