



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JUNE 2023



PC-1024

Lab Name: Govt. PHC, Edayamelur

Lab No: 10596

Constituent Group: Chemistry III

Date of Result Entered: 19/06/2023

PT Item: Lyophilized human serum based

Date of Report Published: 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	284.28 mg/dL	-0.16	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.46 mg/dL	-0.64	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	116.12 mg/dL	0.47	0.24
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :
Govt. PHC, Edayamelur
CLINICAL LABORATORY, SIVAGANGAI District.,
Edayamelur
TAMILNADU630562

Coordinator Contact Details:
Email: cllngc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

தமிழ்நாடு சர்க்கார்
செயல்பாட்டு அமைப்பு
தொலைபேசு: 630 562

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****



CHRISTIAN MEDICAL COLLEGE
 DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - MAY 2023



PC-1024

Lab Name: Govt. PHC, Edayamelur Lab No: 10596

Constituent Group: Chemistry III Date of Result Entered: 20/05/2023

PT Item: Lyophilized human serum based Date of Report Published: 05/06/2023

SL.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	185.18 mg/dL	0.46	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.69 mg/dL	-0.53	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	115.23 mg/dL	0.11	0.27
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

LAB ADDRESS :
 Govt. PHC, Edayamelur
 CLINICAL LABORATORY, SWAGANGAI District,
 Edayamelur
 TAMILNADU 630562

Coordinator Contact Details:
 Email: cmcgo@cmcvellore.ac.in
 Contact Number: 0415-2223102

Paula Christudas
 Dr. Pamela Christudas
 CMC EOAS Coordinator
 Christian Medical College, Vellore

Dr. S. Srinivasan
 Dr. S. Srinivasan
 Sr. Lecturer in Biochemistry
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EOAS reports is confidential
 CMC EOAS does not subcontract any components
 ***** End of Report *****

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Edayamelur
 Lab No 10596
 Month April
 Year 2023
 Constituent Group Chemistry III

Click on the analyte to view Graphical Data

[Details About Robust Analysis](#)
[Detail About Monthly Summary](#)
[Detail about SDI](#)

Date of Result Entered : 20/04/2023

Date of Report Published : 06/05/2023

SLNo	Analyte	Method / Principle Name	Analyser Name	No of Participants	DV	CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	121.23 mg/dL	0.25	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.19 mg/dL	0.17	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	119.98 mg/dL	1.45	0.28
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

Handwritten signature
 மதுரை மருத்துவமனை
 மருத்துவ அலுவலர் 630 562

Links

[Modify Address](#)

[Modify Methods](#)

[Submit Lab Results](#)

[View Lab Monthly Summary](#)