



ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 230728A/N-102	Page	: 1 of 1	
CALIBRATED ON.	: 28 / 07 / 2023	Ref. Std.	: ISO:8655:6	
CALIBRATION DUE.	: 27 / 07 / 2024	C.P. No.	: ACS/CP/08	
SRF No. & DATE	: 230728A/N Dt. 28.07.2023	Phy. Inspection	: O.K.	
ULR No. & ISSUE DATE	: CC34272300000871F Dt. 01.08.2023	Cal. At	: Site / Lab	
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. 34-2-45, A.S. Rama Rao Street, Kasturibai Pet, Vijayawada, Andhra Pradesh-520010.			
ITEM CALIBRATED	: Micro Pipette			
	Make	: Thermo Scientific	Range	: (10 to 100) µl
	L. Count	:	ID No.	: RCL/HEM/PIP-02
	Location	: Haematology	S. No.	: TW08027
ENVIRONMENT CONDITION:	Temp. { 25°C ± 4°C} R.H. {30% to 75%}			
STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/D449927064	BlueBox	BBT/398/JAN/23	05/01/2024
RESULTS :	Mechanical Discipline			
Nominal Capacity (µl)	Graduation (µl)	Nominal Value (µl)	Volume at 20 °C (µl)	Error (µl)
		10	10.10	(-) 0.10
100		50	50.20	(-) 0.20
		100	100.36	(-) 0.36

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 0.1 µl at coverage factor k=1.96.

*U.C. : Under Calibration.

1. Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
2. The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
3. This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
4. The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
5. Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
6. Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

