



ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



Calibration Certificate

| | | | | |
|--|---|---------------------------|-----------------------------|-------------------|
| CERTIFICATE NO. | : 230728A/N-105 | Page | : 1 of 1 | |
| CALIBRATED ON. | : 28 / 07 / 2023 | Ref. Std. | : ISO:8655:6 | |
| CALIBRATION DUE. | : 27 / 07 / 2024 | C.P. No. | : ACS/CP/08 | |
| SRF No. & DATE | : 230728A/N Dt. 28.07.2023 | Phy. Inspection | : O.K. | |
| ULR No. & ISSUE DATE | : CC342723000000874F Dt. 01.08.2023 | Cal. At | : Site / Lab | |
| CALIBRATED FOR | : M/s Redcliffe Lifetech Pvt. Ltd. 34-2-45, A.S. Rama Rao Street, Kasturibai Pet, Vijayawada, Andhra Pradesh-520010 | | | |
| ITEM CALIBRATED | : Micro Pipette | | | |
| | Make | : Thermo Scientific | Range | : (10 to 100) µl |
| | L. Count | : | ID No. | : RCL/BIO/PIP-05 |
| | Location | : Clinical Biochemistry | S. No. | : TW08010 |
| ENVIRONMENT CONDITION: | Temp. { 25°C ± 4°C} R.H. {30% to 75%} | | | |
| STANDARD INSTRUMENTS USED : | | | | |
| INSTT. NAME | ID No. / S. No. | TRACEABILITY | CERTIFICATE No. | DUE DATE |
| Weighing Balance | ACS/DB/02/D449927064 | BlueBox | BBT/398/JAN/23 | 05/01/2024 |
| RESULTS : Mechanical Discipline | | | | |
| Nominal Capacity (µl) | Graduation (µl) | Nominal Value (µl) | Volume at 20 °C (µl) | Error (µl) |
| | | 10 | 10.11 | (-) 0.11 |
| 100 | | 50 | 50.21 | (-) 0.21 |
| | | 100 | 100.40 | (-) 0.40 |

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 0.1 µl at coverage factor k=1.96.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by *Rg*

....End of Calibration Certificate....

