

				CALIB	RATIO	ON CL	ERTIFI	CATE	
CERTIFICATE NO: SBS/CL/23/09307					MEDICAL DEVICES			Page No:1 of 1	
Issue Date				29-07	29-07-2023				
SRF No & Date				SRF/	SRF/23/00276-0003 & 27-07-2023				
Receipt Date				27-07	27-07-2023				
Calibration Date				27-07	27-07-2023				
Calibration Due				26-07	26-07-2024				
Custo	mer Name	& Addr	ess				Act To be a little and a little		
GOVE	RNMENT	PRIM	ARY HEALTH	H CENTRE,		The state of the			
VENC	SALUR, DE	EVANK	OTTAI-63030	3, SIVAGANGAI	DISTRICT				
				Deta	ils of Devi	ce Under C	alibration (D	UC)	
Descr	iption	: ELECTRICAL SAFE ANALYZER(MICRO							00
Range			MULTI		Sr. N	0	and the state of the	190211459	
Resol		:	MULTI		Identi	ification No		NA	
DUC	Condition		SATISFAC	TORY	Local	tion	:	LAB	
	Mark California	REAL PROPERTY.	EN PROPERTY.	Environment	tal Conditi	ons & Calil	oration Proce	dure Details	
Environmental Details Temperature: 25.3°C					Relative Humidity		53% RH		
Calibration Procedure No SBS/CP/MD/14				Calibration done at		ONSITE			
1999	190		HUTTO FEET MARRIES	《	Referen	ce Standar	ds Details		
S.No Description Make			Make/ S	Make/ SI No:		Certificat	e No	Validity	
Electrical Safety Analyser		Rigel Me	Rigel Medical & 44L-1059		TSC/22-23/7400-3		10-08-2023		

ELECTRICAL SAFETY

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resistance	Measured values in MΩ	Uncertainty in % (±)
300	>20MΩ	93	13.92
2	Earth Leakage	Measured values in μA	Uncertainty in % (±)
	<5000µAfor B,BF,CF	121	7.7
3	Enclosure Leakage	Measured values in μA	Uncertainty in % (±)
	<500µAfor B,BF,CF	226	6.8

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- Equipment used for Calibration were calibrated & traceable to National & International Standards.
 Calibrated By

(Calibration Engineer)

K.SATHYAMOORTHY

Chennai 600 032

Technical Manager (C.SHANMUGARAJ)

Chief Executive

Authorised Signatory