

Date : 08/02/2023

## CALIBRATION CERTIFICATE

RAJANS HOSPITAL PVT,LTD  
DIAGNOSTICS CENTER  
Ramanathapuram-623501

Dear Sir/ Madam ,

As desired by you please find the enclosed herewith the calibration report.

1. Serial No : 60370
2. Range of the instrument : Optical density Liner from 0-3.5 A
3. Precision & Accuracy study on EM 360 done on 08/02/2023.

Next Calibration due on or before 07/02/2024

This instrument is manufactured in accordance with the European law in force, in particular to "Electricity Safety" EN 610101 and Electromagnetic Compatibility " EN 61326 Standards.

WAVE LENGTH	GAIN
340	120
376	126
415	137
450	139
480	123
505	129
546	131
570	139
600	141
660	143
700	140
750	132

### TRANSASIA BIO-MEDICALS LTD.

Transasia House, 8 Chandivali Studio Road, Andheri (E), Mumbai - 400 072 Tel.: (022) 40309000, Fax : (022) 2857 3030  
Email : transasia@transasia.co.in Website : www.transasia.co.in

	IDEAL	ACTUAL
Pressure	0.99 bar	1.0 bar
Lamp Voltage	12 V	11.74V
Temperature	37 C	37 C

Photometer Calibration Check	OK
Cuvette Blank Check	OK

The instrument does photometer check before each batch of tests is run. It will give an error message if the photometer test fails and it will not run any samples.

Thanking you and assuring of our prompt attention at all times.

Your faithfully,  
For Transasia-Bio Medicals Ltd,



Sr.Service Engineer

**TRANSASIA BIO-MEDICALS LTD.**

Transasia House, 8 Chandivali Studio Road, Andheri (E), Mumbai - 400 072 Tel.: (022) 40309000, Fax : (022) 2857 3030  
Email : transasia@transasia.co.in Website : www.transasia.co.in

Date: 02/03/2023

**CALIBRATION CERTIFICATE****TO WHOM SOEVER IT MAY CONCERN**

This is to certify that the **Easylyte** analyzer bearing the serial no: 65638CNKC has installed by us at **RAJANS HOSPITAL DIAGNOSTIC CENTRE, RAMANATHAPURAM** is automatic self calibrating electrolyte analyser for estimation of Sodium, Pottasium, Chloride.

The following Diagnostics checks were Carried on 1-MAR-2023

The Next Calibration due is on or before : **1-MAR-2024**

Sodium Calibration Value.	- 58.97	(50-64 mmol/l)
Pottasium Calibration Value	- 57.57	(50-64 mmol/l)
Chloride Calibration Value	- 48.04	(40-54 mmol/l)

Fluid path test : ok  
Pump tube test : ok  
Sample Detector : ok

For TRANSASIA BIO MEDICALS LTD



R. BALAMURUGAN  
SR.SERVICE ENGINEER

**TRANSASIA BIO-MEDICALS LTD.**

Transasia House, 8 Chandivali Studio Road, Andheri (E), Mumbai - 400 072 Tel.: (022) 40309000, Fax : (022) 2857 3030  
Email : [transasia@transasia.co.in](mailto:transasia@transasia.co.in) Website : [www.transasia.co.in](http://www.transasia.co.in)

Date: 08/02/2023


**CALIBRATION CERTIFICATE****TO WHOM SOEVER IT MAY CONCERN**

This is to certify that the **Erba Chem-7** Biochemistry analyzer bearing the serial no: N140277 has installed by us at **RAJANS HOSPITAL, PVT,LTD,DIAGNOSTICS CENTER,Ramanathapuram-623501**

Has been calibrated on 08-02-2023 and verified to control values.

**The Next Calibration Due is on or Before : 07-02-2024**

For TRANSASIA BIO MEDICALS LTD



R. BALAMURUGAN  
Sr.SERVICE ENGINEER

**TRANSASIA BIO-MEDICALS LTD.**

Transasia House, 8 Chandivali Studio Road, Andheri (E), Mumbai - 400 072 Tel.: (022) 40309000, Fax : (022) 2857 3030  
Email : [transasia@transasia.co.in](mailto:transasia@transasia.co.in) Website : [www.transasia.co.in](http://www.transasia.co.in)