

Erba Lachema s.r.o.



Brno, 10.11.2016

To,  
Redcliffe Lifetech Private Limited,  
Vijayawada, Andhra Pradesh - 520010

Instrument Name: ERBA LAURA  
Instrument Serial No: 1061-1-202209

We Erba Lachema SRO with registered office in Czech Republic, Karasek 2219/1D, 621 00 BMO, represented by its Production manager Veronika Pleskova here by confirm as Producer of LAURA instrument that this device is firmly calibrated by the manufacturer and does not require any regular calibration by the user, correct calibration of instrument is controlled by grey strips provided initially together with the instrument.

*P. F.*  
A handwritten signature in blue ink, appearing to be "P. F." followed by a stylized signature.

RNDr. Pavel Suchan Ph.D.

Global Product Manager, Urine Analysis

Erba Lachema s.r.o.

Karásek 1d, Brno 62100, Czech Republic



DEKAPHAN LAURA  
Seq.No: 0008  
URINORM P  
2023.08.21

12:04

.....  
\*BLD 50 Ery/u1  
\*BIL 3 mg/dl  
\*UBG 1 mg/dl  
\*KET 52 mg/dl  
\*GLU 500 mg/dl  
\*PRO 500 mg/dl  
pH 9  
\*NIT POS  
\*LEU TRACE  
SG 1.000

-----  
QC TEST: OK

DEKAPHAN LAURA  
Seq.No: 0007  
URINORM N  
2023.08.21

12:02

.....  
BLD NEG  
BIL NEG  
UBG NORM  
KET NEG  
GLU NEG  
PRO NEG  
pH 6  
NIT NEG  
LEU NEG  
SG 1.025

*M. Loke*

-----  
QC TEST: OK

*Cal Anni*