



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JUNE 2023



PC-1024

Lab Name	MEDRAY CLINICS PVT LTD	Lab No	17475
Constituent Group	HbA1c	Date of Result Entered :	16/06/2023
PT item	Lyophilized human whole blood based	Date of Report Published :	04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	HbA1c	HPLC-Ion Exchange	Any Analyser	164	6.34	8.79	0.56	6.1 %	-0.43	0.09

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

MEDRAY CLINICS PVT LTD
NO.962, 12TH MAIN ROAD, NEAR RELIANCE DIGITAL HALL 2ND STAGE, INDIRA NAGAR
BANGALORE
KARNATAKA560008

Pamela Christudoss

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****