

Erba Lachema s.r.o.



Brno, 10.11.2016

To,  
Redcliffe Lifetech Private Limited,  
Vishakapatnam, Andhra Pradesh - 530002

Instrument Name: ERBA LAURA  
Instrument Serial No: 1062-1-202209

We Erba Lachema SRO with registered office in Czech Republic, Karasek 2219/1D, 621 00 BMO, represented by its Production manager Veronika Pleskova here by confirm as Producer of LAURA instrument that this device is firmly calibrated by the manufacturer and does not require any regular calibration by the user, correct calibration of instrument is controlled by grey strips provided initially together with the instrument.

*P.S.*  
A handwritten signature in blue ink, appearing to be "P.S.", followed by a larger, more complex signature.

RNDr. Pavel Suchan Ph.D.

Global Product Manager, Urine Analysis

Erba Lachema s.r.o.

Karásek 1d, Brno 62100, Czech Republic



Doherty  
T. 29  
23/08/23

DEKAPHAN LAURA

Seq.No: 0001

URINORM N

2023.08.23

09:35

.....

BLD	NEG
BIL	NEG
UBG	NORM
KET	NEG
GLU	NEG
PRO	NEG
pH	6
NIT	NEG
LEU	NEG
SG	1.025

QC TEST: OK

DEKAPHAN LAURA

Seq.No: 0002

URINORM P

2023.08.23

09:37

.....

BLD	50	Ery/u1
BIL	3	mg/dl
UBG	3	mg/dl
KET	52	mg/dl
GLU	500	mg/dl
PRO	100	mg/dl
pH	9	
NIT	POS	
LEU	75	Leu/u1
SG	1.000	

QC TEST: OK