


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 230712A/N-115	Page	: 1 of 1	
CALIBRATED ON.	: 14 / 07 / 2023	Ref. Std.	: IS:12508:1988	
CALIBRATION DUE.	: 13 / 07 / 2024	C.P. No.	: ACS/CP/10	
SRF No. & DATE	: 230712A/N Dt. 12.07.2023	Phy. Inspection	: O.K.	
ULR No. & ISSUE DATE	: CC342722000000790F Dt. 15.07.2023	Cal. At	: Site / Lab	
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. First Floor 14-37-38, Krishna Nagar, Maharanieta, Andhra Pradesh-530002			
ITEM CALIBRATED	: Centrifuge			
	Make	: Remi	Range	: (0 to 6000) RPM
	L. Count	:	ID No.	: RCL/CP/Cent-02
	Location	: Clinical Pathology	S. No.	: ZIEN-24020
ENVIRONMENT CONDITION:	Temp. { 25°C ± 3°C } R.H. { 40% to 60% }			
STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Tachometer	ACS/DTM/01/ 211-0155V	Godrej	C-220912-6-1	12/09/2023
RESULTS	: Mechanical Discipline			
	* U. C. INSTT. READS (RPM)	Standard Measured (RPM)	Error (RPM)	
	1000	1001.22	- 1.22	
	2000	2002.24	- 2.24	
	4000	4005.38	- 5.38	
	6000	6006.22	- 6.22	
	(average of readings)			

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 2.50 rpm of reading at coverage Factor $k=1.96$.

*U.C. : Under Calibration.

1. Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
2. The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
3. This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
4. The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
5. Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
6. Confirmity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

