


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO. : 230712A/N-105	Page : 1 of 1
CALIBRATED ON. : 12 / 07 / 2023	Ref. Std. : ISO:8655:6
CALIBRATION DUE. : 11 / 07 / 2024	C.P. No. : ACS/CP/08
SRF No. & DATE : 230712A/N Dt. 12.07.2023	Phy. Inspection : O.K.
ULR No. & ISSUE DATE : CC342722000000780F Dt. 15.07.2023	Cal. At : Site / Lab
CALIBRATED FOR : M/s Redcliffe Lifetech Pvt. Ltd. First Floor 14-37-38, Krishna Nagar, Maharanipeta, Andhra Pradesh-530002	

ITEM CALIBRATED : Micro Pipette	
Make : Dragon Lab	Range : (10 to 100) µl
L. Count :	ID No. : RCL/BIO/PIP-05
Location : Clinical Biochemistry	S. No. : TW06252

ENVIRONMENT CONDITION: Temp. { 25°C ± 3°C } R.H. {40% to 60%}

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/ D449927064	BlueBox	BBT/398/JAN/23	05/01/2024

RESULTS : Mechanical Discipline				
Nominal Capacity (µl)	Graduation (µl)	Nominal Value (µl)	Volume at 20 °C (µl)	Error (µl)
100		10	10.12	(-) 0.12
		50	50.25	(-) 0.25
		100	100.45	(-) 0.45

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 0.6 µl at coverage factor k=2.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.

Prepared by *Rej*

....End of Calibration Certificate....

