

CALIBRATION REPORT

Centrifuge Table top



Report No:	CAL-AR-2022-01-0008	Date: 10-01-2022	Page 01 of 02
------------	---------------------	------------------	---------------

Customer Details			
Health Facility	Bakin Pertin GH, Pasighat		
State	Arunachal Pradesh	District	East Siang

Medical Equipment Details			
ME Number	ARZ1EASDH0150010	Manufacturer	Remi
ME Name	Centrifuges, Tabletop (Centrifuge)	Model	NA
Department	Laboratory-Sti Lab	Serial No	NA
Calibrated at	On Site	Condition of ME	Good
Date of Calibration	27-12-2021'	Due for Calibration	26-12-2022'

Environmental Conditions			
Temperature in °C	20°C	Humidity in % Rh	57%

Details of Master Equipment Used for Calibration						
No	Equipment ID	Description	Manufacturer	Model	Serial No.	Calibration Validity
1	TRI-TE-0015	Digital Tachometer	AMBROBE	TACH-10	16060021	17-09-2022'

Performance Results						
No	Parameters	Unit	Set Value	Measured Value	Acceptable Range	Remarks
1	Speed	rpm	Set 1	1520	1425 to 1575	Pass
2	Speed	rpm	Set 2	2530	2375 to 2625	Pass
3	Speed	rpm	Set 3	3550	3325 to 3675	Pass

CALIBRATION REPORT

Centrifuge Table top



Conditions of Issue of report

1. The Standards used are traceable to National standards.
2. This report should not be reproduced except in full without written approval.

Calibration Status

Passed Failed

Calibrated By

A handwritten signature in blue ink, appearing to be "Shashi Shekhar".

Shashi Shekhar

Authorised Signatory

A handwritten signature in blue ink, appearing to be "S Palaniswamy".

S Palaniswamy