

CALIBRATION REPORT

Refrigerators, Laboratory(Lab refrigerator)



Report No:	CAL-AR-2022-01-0007	Date: 07-01-2022	Page 01 of 02
------------	---------------------	------------------	---------------

Customer Details			
Health Facility	Bakin Pertin GH, Pasighat		
State	Arunachal pradesh	District	East Siang

Medical Equipment Details			
ME Number	ARZ1EASDH0150009	Manufacturer	LG
ME Name	Refrigerators, Laboratory(Lab refrigerator)	Model	INTELLO COOI
Department	Laboratory-Sti Lab	Serial No	NA
Calibrated at	On Site	Condition of ME	Good
Date of Calibration	27-12-2021	Due for Calibration	26-12-2022

Environmental Conditions			
Temperature in °C	20°C	Humidity in % Rh	59%

Details of Master Equipment Used for Calibration						
No	Equipment ID	Description	Manufacturer	Model	Serial No.	Calibration Validity
1	TRI-TE-0187	DIGITAL THERMOMETER	METRAVI	DTM-900	271926/1	17-09-2022

Performance Results						
No	Parameter	Set Temp	Display Temp	Measured Value	Acceptable Range	Remarks
1	Temperature	2°C	°C	2	1.5 to 2.5°C	Pass
2	Temperature	4°C	°C	4.1	3.5 to 4.5°C	Pass
3	Temperature	6°C	°C	6.2	5.5 to 6.5°C	Pass

CALIBRATION REPORT

Refrigerators, Laboratory(Lab refrigerator)



Conditions of Issue of report

1. The Standards used are traceable to National standards.
2. This report should not be reproduced except in full without written approval from Trimed Solutions India Private Limited.

Calibration Status

Passed Failed

Calibrated By

A handwritten signature in blue ink, appearing to be "Shashi Shekhar".

Shashi Shekhar

Authorised Signatory

A handwritten signature in blue ink, appearing to be "S Palaniswamy".

S Palaniswamy