



Customer: Guru Greencross Pathology Laboratory

Equipment: BS 240 FULLY AUTO BIOCHEMISTRY ANALYZER **Date :** 05/05/2023

Title: Instrument Calibration Certificate **Certificate No.:** TMQ/H/GGPL/23-24

Instrument Detail

Equipment	MODEL	SERIAL NO.	SOFTWARE VERSION
FULLY AUTO BIOCHEMISTRY ANALYZER	BS240	YX-18004508	V24.00.07

Calibration Procedure

No.	Unit	Parameter	Calibrated & Verified
1.	Sample & Reagent Carousel	Sample/Reagent Carousel Outer Ring Aspiration Stop Position	<input type="checkbox"/>
		Sample/Reagent Carousel Middle Ring Aspiration Stop Position	<input type="checkbox"/>
		Sample/Reagent Carousel Inner Ring Aspiration Stop Position	<input type="checkbox"/>
		Outer Ring Barcode Scanning	<input type="checkbox"/>
		inner Ring Barcode Scanning	<input type="checkbox"/>
		Probe Vertical/Horizontal Position on Wash Well	<input type="checkbox"/>
		Horizontal Aspirate Position on Sample/Reagent Carousel Outer Ring	<input type="checkbox"/>
		Horizontal Aspirate Position on Sample/Reagent Carousel Middle Ring	<input type="checkbox"/>
		Horizontal Aspirate Position on Sample/Reagent Carousel Inner Ring	<input type="checkbox"/>
2.	Mixer	Mixer Horizontal Wash Position	<input type="checkbox"/>
		Sample Mixing Horizontal Position	<input type="checkbox"/>
		Sample Mixing Vertical Position	<input type="checkbox"/>

Customer: Guru Greencross Pathology Laboratory

Equipment: BS 240 FULLY AUTO BIOCHEMISTRY ANALYZER

Date : 05/05/2023

Title: Instrument Calibration Certificate

Certificate No.: TMQ/H/GGPL/23-24

3.	Reaction Carousel	Photo Electricity Gain Channel 340	<input type="checkbox"/>
		Photo Electricity Gain Channel 405	<input type="checkbox"/>
		Photo Electricity Gain Channel 450	<input type="checkbox"/>
		Photo Electricity Gain Channel 510	<input type="checkbox"/>
		Photo Electricity Gain Channel 546	<input type="checkbox"/>
		Photo Electricity Gain Channel 578	<input type="checkbox"/>
		Photo Electricity Gain Channel 630	<input type="checkbox"/>
		Photo Electricity Gain Channel 670	<input type="checkbox"/>
4.	Temperature Control	Reaction Carousel Sensor 1 A	<input type="checkbox"/>
		Reagent Carousel Sensor 1 B	<input type="checkbox"/>
		Pre-Heater Temperature 1 C	<input type="checkbox"/>
5.	Cuvette Automatic Wash	Reaction Carousel dispensing position	<input type="checkbox"/>
		Reaction Carousel washing position.	<input type="checkbox"/>
		Fluidic Prime Complete	<input type="checkbox"/>



Customer: Guru Greencross Pathology Laboratory

Equipment: BS 240 FULLY AUTO BIOCHEMISTRY ANALYZER

Date : 05/05/2023

Title: Instrument Calibration Certificate

Certificate No.: TMQ/H/GGPL/23-24

Calibrated By

Name: Hardik Patel

Sign & Date:

A circular stamp with the text "TANISHH MEDI QUIP" around the top edge and "SURAT" in the center. Below the stamp is a handwritten signature in black ink that reads "Hardik Patel".

Hardik Patel

Designation: Service Engineer