



CIN-U51507DL1996PLC079544



HD Consortium India Limited

Regd. Office : A-11, First Floor, Okhla Industrial Area,
Phase-I, New Delhi-110020
Phone : +91-11-40527276
E-mail : sales@hdcindia.com, services@hdcindia.com
Website : www.hdcindia.com

CALIBRATION REPORT

Dated :- 01 April, 2023

**To,
Guru Greencross Pathology Laboratory Medipolis
A-2, 4th floor Medipolis, New Doctor House,
Deesa Road, Palanpur – 385001**

Calibration Done Date :- **01 April, 2023**
Calibration Due Date :- **31 March, 2024**

Dear Sir,

This is to certify that your **Electrolyte Analyzer, Model :- HDC-Lyte** bearing Sl. No. **HDCL-3493** is Serviced and Calibrated.

It is working fine and giving satisfactory result.

<u>Parameters</u>	<u>Slope</u>
Na+	60.47
K+	61.54
iCa+	31.45

Thanking you and assuring you of our best services and attention always.

Your Sincerely,

Engineer's Name & Signature



Authorized Signatory

Customer's Name & Signature

Authorized Signatory

YOUR PARTNER IN PATIENT-CARE