

CENTRIFUGE Calibration / Verification Record Sheet			
Location: ICTC,SVIMS, Tirupati		Range : Up to 3500RPM	
Equipment ID: ICTC/SVIMS/TPT/CF/01		SLNo. : Z1CN-13054	
Calibration Procedure Reference: In-house verification of centrifuge using certified standard calibrated by Tachometer			
Certified centrifuge ID: SRL/HIVTL/Tachometer		Calibration date : 21.06.2023	
Certified centrifuge calibration date : 30.03.2023		Calibration due date : 20.06.2024	
Verification Done by: G.Archana, Technical Officer, SRL, SVMC, Tirupati			
S.NO	STANDARD READING(RPM)	UUC READING (RPM)	ERROR (RPM)
1	499	500	1.0
2	998.5	1000	1.5
3	1997.4	2000	2.6
4	2996.7	3000	3.3
5	3496.7	3500	3.3
Accept/ Reject: Accepted since average difference is within +/- 100RPM			
Final Interpretation: Centrifuge is in good working condition			
Action Taken: - NA			
Remark: - NA			
Accept/ Reject: Accepted since average difference is within +/- 100RPM			

Archana

Signature

Technical Officer
APSACS (SRL)

Department of Microbiology
S.V. Medical College/SVRRGG
TIRUPATI - 517 507